



Calibration Services
 5750 SW Arctic Drive
 Beaverton, Oregon 97005
 USA



Calibration Cert #2527.01

Calibration Report

Accredited Calibration

Report Number: APx525-XXXXX-XXXXXX-X

Model: APx525
 Serial Number: APX2-XXXXX

Data Type: AS RECEIVED
 Date of Cal: 4-Apr-2014

Program:
 APxCalibration.exe 3.43

Internal Module Status and Data			
AP Name	Description	Serial No.	Revision
BRIO	Backplane module	XXXX	108
BSOL	Clock module	XXXX	103
BMEG	Analog Output module	XXXX	006
BOTA	Analog Input module	XXXX	007
BAES	Digital I/O module	XXXX	102
BSAT	Option AG52	XXXX	100

Explanatory notes to the last three columns of the calibration report

"MU" - The column labeled "MU" lists the expanded measurement uncertainties derived from equipment specifications, repeatability data, and other significant sources. These are stated at a minimum confidence of 95% using a coverage factor k=2 (except as appropriate) following the recommendations in ISO/IEC 98-3 *Guide to the expression of uncertainty in measurement (GUM:1995)*, BIPM JCGM 100:2008, and NIST Technical Note 1297.

"TUR" - The column labeled "TUR" lists the test uncertainty ratio calculated by dividing the lesser of the lower and upper reading tolerances by the 95% expanded measurement uncertainty. An entry of "na" indicates [1] the specified limits are one-sided, or [2] the performance characteristic is not accredited.

"Result" - The column labeled "Result" lists color-coded assessments that the observed characteristic is within its specified limits of performance. There are three possible indications:

pass -- The *READING* is within the specified upper and lower limits reduced by guard-bands equal to the 95% expanded measurement uncertainty. The probability or risk of false acceptance is very low, typically <0.2%.

uncertain -- The *READING* is within the specified upper and lower limits, but it is close to one of the limits by an amount that is less than the 95% expanded measurement uncertainty. The probability or risk of false acceptance is elevated.

>> FAIL << -- The *READING* is outside of the specified limit range.

Accredited measurements listed in the following pages correlate to Audio Precision's Scope of Accreditation as noted:

- note 1 - Frequency Measurement
- note 2 - AC Voltage Measurement
- note 3 - AC Flatness Measurement
- note 4 - Resistance Measurement
- note 5 - DC Voltage Measurement
- note 6 - AC Voltage Source for testing AC Measurement Equipment
- note 7 - DC Voltage Source for testing DC Measurement Equipment

This report is valid only when accompanied by a signed Certificate of Calibration.

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG GENERATOR CHARACTERISTICS							
[1] Sine Frequency Accuracy (Hz) - note 1							
	10 kHz	9999.9700	10000.0013	10000.0300	0.0054	5.6	pass
[2] Sine Amplitude Accuracy, 1 kHz (Volts, mVolts) - note 2							
Channel 1	13.33 V	13.2840	13.3325	13.3761	0.0058	8.0	pass
	5.000 V	4.9828	5.0047	5.0173	0.0013	≥10	pass
	3.000 V	2.98966	3.00104	3.01038	0.00096	≥10	pass
	1.500 V	1.49483	1.50064	1.50519	0.00048	≥10	pass
	1.000 V	0.99655	1.00008	1.00346	0.00034	≥10	pass
	100 mV	99.655	100.017	100.346	0.037	9.5	pass
	12 mV	11.9586	12.0065	12.0415	0.0061	6.9	pass
Channel 2	13.33 V	13.2840	13.3328	13.3761	0.0058	8.0	pass
	5.000 V	4.9828	5.0041	5.0173	0.0013	≥10	pass
	3.000 V	2.98966	3.00132	3.01038	0.00096	≥10	pass
	1.500 V	1.49483	1.50075	1.50519	0.00048	≥10	pass
	1.000 V	0.99655	1.00016	1.00346	0.00034	≥10	pass
	100 mV	99.655	100.027	100.346	0.037	9.5	pass
	12 mV	11.9586	12.0066	12.0415	0.0061	6.9	pass
[3] Sine Flatness, 1.15 Vrms (dB) - note 3							
Channel 1	10 Hz	-0.0080	-0.0001	0.0080	0.0016	5.0	pass
	20 Hz	-0.0080	-0.0001	0.0080	0.0015	5.3	pass
	20 kHz	-0.0080	-0.0039	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0056	0.0300	0.0038	7.9	pass
	80 kHz	-0.1000	-0.0024	0.1000	0.0075	≥10	pass
Channel 2	10 Hz	-0.0080	-0.0001	0.0080	0.0016	5.0	pass
	20 Hz	-0.0080	0.0000	0.0080	0.0015	5.3	pass
	20 kHz	-0.0080	-0.0044	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0065	0.0300	0.0038	7.9	pass
	80 kHz	-0.1000	-0.0049	0.1000	0.0075	≥10	pass
[4] DC Offset (mV), Unbal - note 5							
Channel 1	10.0 V	-25.10	3.86	25.10	0.40	≥10	pass
	1.000 V	-2.600	0.337	2.600	0.046	≥10	pass
	100 mV	-0.350	0.049	0.350	0.024	≥10	pass
	10 mV	-0.125	0.018	0.125	0.024	5.2	pass
Channel 2	10.0 V	-25.10	4.38	25.10	0.40	≥10	pass
	1.000 V	-2.600	0.449	2.600	0.046	≥10	pass
	100 mV	-0.350	0.055	0.350	0.024	≥10	pass
	10 mV	-0.125	0.034	0.125	0.024	5.2	pass
[5] Source Resistance Accuracy (Ω) - note 4							
Channel 1 Unbalanced	20 Ω	19.600	20.064	20.400	0.010	≥10	pass
	50 Ω	49.250	50.046	50.750	0.020	≥10	pass
	75 Ω	74.100	74.921	75.900	0.020	≥10	pass
	100 Ω	99.000	99.805	101.000	0.040	≥10	pass
	600 Ω	594.00	598.63	606.00	0.16	≥10	pass
Channel 1 Balanced	40 Ω	39.400	39.874	40.600	0.020	≥10	pass
	100 Ω	99.000	99.834	101.000	0.040	≥10	pass
	150 Ω	148.500	149.592	151.500	0.060	≥10	pass
	200 Ω	198.000	199.345	202.000	0.060	≥10	pass
	600 Ω	594.00	598.44	606.00	0.16	≥10	pass
Channel 2 Unbalanced	20 Ω	19.600	20.098	20.400	0.010	≥10	pass
	50 Ω	49.250	50.070	50.750	0.020	≥10	pass
	75 Ω	74.100	74.956	75.900	0.020	≥10	pass
	100 Ω	99.000	99.828	101.000	0.040	≥10	pass
	600 Ω	594.00	598.77	606.00	0.16	≥10	pass
Channel 2 Balanced	40 Ω	39.400	39.913	40.600	0.020	≥10	pass
	100 Ω	99.000	99.867	101.000	0.040	≥10	pass
	150 Ω	148.500	149.638	151.500	0.060	≥10	pass
	200 Ω	198.000	199.377	202.000	0.060	≥10	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
	600 Ω	594.00	598.60	606.00	0.16	≥ 10	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER CHARACTERISTICS							
[6] Input Termination Accuracy (Ω) - note 4							
Channel 1	300 Ω	297.000	301.597	303.000	0.080	≥ 10	pass
	600 Ω	594.00	603.40	606.00	0.16	≥ 10	pass
Channel 2	300 Ω	297.000	301.774	303.000	0.080	≥ 10	pass
	600 Ω	594.00	602.89	606.00	0.16	≥ 10	pass
[7] DC Measurement Accuracy (Volts, mVolts) - note 7							
Channel 1	+160 V	158.720	159.906	161.280	0.064	≥ 10	pass
	+100 V	99.200	99.943	100.800	0.040	≥ 10	pass
	+30 V	29.758	29.984	30.242	0.012	≥ 10	pass
	+10 V	9.9200	9.9943	10.0800	0.0040	≥ 10	pass
	+3 V	2.9758	2.9982	3.0242	0.0012	≥ 10	pass
	+1 V	0.99200	0.99939	1.00800	0.00040	≥ 10	pass
	+300 mV	297.10	299.83	302.90	0.12	≥ 10	pass
	0 mV	-0.800	-0.013	0.800	0.046	≥ 10	pass
	-300 mV	-302.90	-299.86	-297.10	0.12	≥ 10	pass
	-1 V	-1.00800	-0.99938	-0.99200	0.00040	≥ 10	pass
	-3 V	-3.0242	-2.9981	-2.9758	0.0012	≥ 10	pass
	-10 V	-10.0800	-9.9940	-9.9200	0.0040	≥ 10	pass
	-30 V	-30.242	-29.981	-29.758	0.012	≥ 10	pass
	-100V	-100.800	-99.938	-99.200	0.040	≥ 10	pass
-160 V	-161.280	-159.900	-158.720	0.064	≥ 10	pass	
Channel 2	+160 V	158.720	159.897	161.280	0.064	≥ 10	pass
	+100 V	99.200	99.935	100.800	0.040	≥ 10	pass
	+30 V	29.758	29.980	30.242	0.012	≥ 10	pass
	+10 V	9.9200	9.9936	10.0800	0.0040	≥ 10	pass
	+3 V	2.9758	2.9981	3.0242	0.0012	≥ 10	pass
	+1 V	0.99200	0.99930	1.00800	0.00040	≥ 10	pass
	+300 mV	297.10	299.76	302.90	0.12	≥ 10	pass
	0 mV	-0.800	-0.069	0.800	0.046	≥ 10	pass
	-300 mV	-302.90	-299.90	-297.10	0.12	≥ 10	pass
	-1 V	-1.00800	-0.99939	-0.99200	0.00040	≥ 10	pass
	-3 V	-3.0242	-2.9980	-2.9758	0.0012	≥ 10	pass
	-10 V	-10.0800	-9.9937	-9.9200	0.0040	≥ 10	pass
	-30 V	-30.242	-29.983	-29.758	0.012	≥ 10	pass
	-100V	-100.800	-99.939	-99.200	0.040	≥ 10	pass
-160 V	-161.280	-159.903	-158.720	0.064	≥ 10	pass	
[8] Input Common Mode Rejection (mV) - note 6							
Channel 1 (5V CM signal)	3.2V range, 200 Hz	0	0.017	0.500	0.070	na	pass
	3.2V range, 5 kHz	0	0.050	0.500	0.082	na	pass
	3.2V range, 20 kHz	0	0.195	1.256	0.180	na	pass
	10V range, 20 kHz	0	0.88	15.81	0.22	na	pass
	100V range, 20 kHz	0	2.11	28.12	0.27	na	pass
Channel 2 (5V CM signal)	3.2V range, 200 Hz	0	0.017	0.500	0.070	na	pass
	3.2V range, 5 kHz	0	0.029	0.500	0.082	na	pass
	3.2V range, 20 kHz	0	0.109	1.256	0.180	na	pass
	10V range, 20 kHz	0	2.11	15.81	0.22	na	pass
	100V range, 20 kHz	0	2.04	28.12	0.27	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[9] Level Meter AC Accuracy, 1 kHz (Volts, mVolts) - note 6							
Channel 1	160 V	159.448	159.900	160.554	0.059	9.5	pass
	90 V	89.690	89.943	90.311	0.033	9.5	pass
	30 V	29.897	29.982	30.104	0.011	9.5	pass
	9 V	8.9690	8.9945	9.0311	0.0032	9.8	pass
	3 V	2.9897	2.9982	3.0104	0.0011	9.5	pass
	0.9 V	0.89690	0.89945	0.90311	0.00032	9.8	pass
	300 mV	298.97	299.85	301.04	0.11	9.5	pass
	5 mV	4.9828	4.9969	5.0173	0.0022	8.0	pass
Channel 2	160 V	159.45	159.90	160.55	0.059	9.5	pass
	90 V	89.690	89.940	90.311	0.033	9.5	pass
	30 V	29.897	29.980	30.104	0.011	9.5	pass
	9 V	8.9690	8.9940	9.0311	0.0032	9.8	pass
	3 V	2.9897	2.9981	3.0104	0.0011	9.5	pass
	0.9 V	0.89690	0.89941	0.90311	0.00032	9.8	pass
	300 mV	298.97	299.84	301.04	0.11	9.5	pass
	5 mV	4.9828	4.9967	5.0173	0.0022	8.0	pass
[10] Level Meter AC Flatness, 1.15 Vrms (dB) - note 6							
Channel 1	10 Hz	-0.0080	-0.0014	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0003	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0021	0.0080	0.0023	3.5	pass
	50 kHz	-0.0300	0.0089	0.0300	0.0030	≥10	pass
	80 kHz	-0.1000	-0.0112	0.1000	0.0059	≥10	pass
Channel 2	10 Hz	-0.0080	-0.0013	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0003	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0019	0.0080	0.0023	3.5	pass
	50 kHz	-0.0300	0.0069	0.0300	0.0030	≥10	pass
	80 kHz	-0.1000	-0.0138	0.1000	0.0059	≥10	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[11] Phase Measurement Offset (Deg) - non-accredited, self test							
Ch1 - Ch 2	50Hz, AC couple	-0.200	0.008	0.200	0.001	na	pass
	200Hz, DC couple	-0.200	0.000	0.200	0.001	na	pass
	5kHz, DC couple	-0.200	0.001	0.200	0.001	na	pass
	20kHz, DC couple	-0.800	0.004	0.800	0.002	na	pass
	50kHz, DC couple	-2.000	0.010	2.000	0.004	na	pass
[12] Frequency Measurement Accuracy (uHz/Hz) - note 1							
	10 kHz	-3.00	-0.13	3.00	0.54	5.6	pass
[13] Input Residual Crosstalk at 20 kHz (dB) - note 6							
Ch2 into Ch 1		-999	-148.8	-140.0	4.0	na	pass
Ch 1 into Ch 2		-999	-149.0	-140.0	4.0	na	pass
OPTION AG52 CHARACTERISTICS							
[14] Squarewave Amplitude Accuracy (Volts) - note 2							
	100 Hz, 1 Vpk	0.98855	0.99990	1.01158	0.00073	≥10	pass
[15] Squarewave Risetime (usec)- non-accredited, oscilloscope referenced							
	5 kHz, 1 Vpk	0.00	1.56	2.00	0.06	na	pass
[16] Squarewave Even Harmonic Content (dB) - non-accredited, self test							
	20 Hz, 1 Vpk	-999	-112.6	-100.0	4.0	na	pass
	5 kHz, 1 Vpk	-999	-108.1	-100.0	1.0	na	pass
	20 kHz, 1 Vpk	-999	-98.7	-90.0	1.0	na	pass
[17] DIM Amplitude Accuracy (Volts) - note 2							
	DIM100, 1Vpk	0.79579	0.80465	0.81432	0.00059	≥10	pass
[18] Residual IMD, DIM100 (%) - non-accredited, self-test							
Balanced, 21.2Vrms	Ch 1	0.00000%	0.00078%	0.00180%	0.00020%	na	pass
	Ch 2	0.00000%	0.00077%	0.00180%	0.00020%	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result	
NON-ACCREDITED CHARACTERISTICS								
[19] Sine THD+N (dB) - non-accredited, self-test								
Channel 1	10 Hz, 21.2V, 20k BW	-999	-110.4	-105.0	0.8	na	pass	
	Balanced	kHz, 21.2V, 20k BW	-999	-111.4	-105.0	0.8	na	pass
Unbalanced	5 kHz, 21.2V, 20k BW	-999	-111.4	-105.0	0.8	na	pass	
	20 kHz, 21.2V, 20k BW	-999	-111.7	-105.0	1.0	na	pass	
	20 Hz, 2.5V, 20k BW	-999	-110.9	-105.0	0.8	na	pass	
	1 kHz, 2.5V, 20k BW	-999	-113.0	-105.0	0.8	na	pass	
	20 kHz, 2.5V, 20k BW	-999	-112.7	-105.0	1.0	na	pass	
	1 kHz, 2.5V 40k BW	-999	-109.5	-100.0	0.8	na	pass	
	10 kHz, 2.5V, 40k BW	-999	-107.5	-100.0	0.8	na	pass	
	1 kHz, 2.5V 80k BW	-999	-102.4	-92.0	0.8	na	pass	
	20 kHz, 2.5V, 80k BW	-999	-99.9	-92.0	0.8	na	pass	
	Channel 2	10 Hz, 21.2V, 20k BW	-999	-110.3	-105.0	0.8	na	pass
Balanced	kHz, 21.2V, 20k BW	-999	-111.1	-105.0	0.8	na	pass	
	5 kHz, 21.2V, 20k BW	-999	-111.0	-105.0	0.8	na	pass	
Unbalanced	20 kHz, 21.2V, 20k BW	-999	-111.9	-105.0	1.0	na	pass	
	20 Hz, 2.5V, 20k BW	-999	-110.7	-105.0	0.8	na	pass	
	1 kHz, 2.5V, 20k BW	-999	-112.9	-105.0	0.8	na	pass	
	20 kHz, 2.5V, 20k BW	-999	-113.1	-105.0	1.0	na	pass	
	1 kHz, 2.5V 40k BW	-999	-109.6	-100.0	0.8	na	pass	
	10 kHz, 2.5V, 40k BW	-999	-107.6	-100.0	0.8	na	pass	
	1 kHz, 2.5V 80k BW	-999	-102.4	-92.0	0.8	na	pass	
	20 kHz, 2.5V, 80k BW	-999	-99.9	-92.0	0.8	na	pass	
	[20] Residual Crosstalk (dB), Output Related, 20kHz - non-accredited, self-test							
	Unbalanced, 10.6V, 20 kHz.	2 into 1	-999	-153.6	-130.0	2.0	na	pass
1 into 2		-999	-153.4	-130.0	2.0	na	pass	
Balanced, 21.2V, 20 kHz.	2 into 1	-999	-163.7	-130.0	4.0	na	pass	
	1 into 2	-999	-163.4	-130.0	4.0	na	pass	
[21] Residual Noise (uVolts) - non-accredited, self-test								
Balanced Input, inputs shorted.	Ch 1, 20 kHz BW	0	1.06	1.30	0.06	na	pass	
	Ch 2, 20 kHz BW	0	1.06	1.30	0.06	na	pass	
[22] Residual SMPTE IMD (%), 4:1, 60Hz:7kHz - non-accredited, self-test								
Balanced, 21.2 Vrms.	Ch 1	0%	0.00051%	0.00180%	0.00020%	na	pass	
	Ch 2	0%	0.00058%	0.00180%	0.00020%	na	pass	
[23] Residual MOD IMD (%), 4:1, 60Hz:7kHz - non-accredited, self-test								
Balanced, 21.2 Vrms.	Ch 1	0%	0.00034%	0.00180%	0.00020%	na	pass	
	Ch 2	0%	0.00046%	0.00180%	0.00020%	na	pass	
[24] Residual DFD IMD (%), mean 19.5kHz, diff 1kHz - non-accredited, self-test								
Balanced, 21.2 Vrms.	Ch 1	0%	0.00013%	0.00050%	0.00010%	na	pass	
	Ch 2	0%	0.00012%	0.00050%	0.00010%	na	pass	
[25] Digital Output Amplitude Accuracy (Volts) - non-accredited								
Unbal, consumer	0.5 Vpp	0.450	0.496	0.550	0.006	na	pass	
Unbal, professional	1.0 Vpp	0.900	0.979	1.100	0.012	na	pass	
Balanced	5.0 Vpp	4.500	5.076	5.500	0.059	na	pass	

END OF REPORT