



Calibration Services
 9290 SW Nimbus Ave
 Beaverton, Oregon 97008
 USA



Calibration Cert #2527.01

Calibration Report

Accredited Calibration

Report Number: XXXXXXXXXX-XXXXXX-X

Model: APx515 **Data Type:** AS SHIPPED, NEW **Program:**
Serial Number: XXXXXXXXXX **Date of Cal:** 14-Apr-2026 APxCalibration.exe 24.01

Internal Module Status and Data			
AP Name	Description	Serial No.	Revision
BXAN	Main Board	71393	702
Advanced Measurement Software License SW-AML			installed

Explanatory notes to the last three columns of the calibration report

"MU" - The column labeled "MU" lists the expanded measurement uncertainties derived from equipment specifications, repeatability data, and other significant sources. These are stated at a minimum confidence of 95% using a coverage factor k=2 (except as appropriate) following the recommendations in ISO/IEC 98-3 *Guide to the expression of uncertainty in measurement (GUM:1995)*, BIPM JCGM 100:2008, and NIST Technical Note 1297.

"TUR" - The column labeled "TUR" lists the test uncertainty ratio calculated by dividing the lesser of the lower and upper reading tolerances by the 95% expanded measurement uncertainty. An entry of "na" indicates [1] the specified limits are one-sided, or [2] the performance characteristic is not accredited.

"Result" - The column labeled "Result" lists color-coded assessments that the observed characteristic is within its specified limits of performance. There are three possible indications:

pass -- The **READING** is within the specified upper and lower limits reduced by guard-bands equal to the 95% expanded measurement uncertainty. The probability or risk of false acceptance is very low, typically <0.2%.

uncertain -- The **READING** is within the specified upper and lower limits, but it is close to one of the limits by an amount that is less than the 95% expanded measurement uncertainty. The probability or risk of false acceptance is elevated.

>> FAIL << -- The **READING** is outside of the specified limit range.

Accredited measurements listed in the following pages correlate to Audio Precision's Scope of Accreditation as noted:

- note 1 - Frequency Measurement
- note 2 - AC Voltage Measurement
- note 3 - AC Flatness Measurement
- note 4 - Resistance Measurement
- note 5 - DC Voltage Measurement
- note 6 - AC Voltage Source and AC Flatness Source for testing AC Measurement Equipment
- note 7 - DC Voltage Source for testing DC Measurement Equipment

This report is valid only when accompanied by a signed Certificate of Calibration.

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result	
ANALOG GENERATOR CHARACTERISTICS								
[1] Sine Frequency Accuracy (Hz) - note 1								
	10 kHz	9999.9700	10000.0037	10000.0300	0.0063	4.8	pass	
[2] Sine Amplitude Accuracy, 1 kHz (Volts, mVolts) - note 2								
Channel 1	8.00 V	7.9541	8.0017	8.0462	0.0022	≥10	pass	
	5.000 V	4.9713	5.0008	5.0289	0.0014	≥10	pass	
	3.000 V	2.9828	3.0020	3.0173	0.0011	≥10	pass	
	1.500 V	1.49139	1.50078	1.50866	0.00054	≥10	pass	
	1.000 V	0.99426	1.00018	1.00577	0.00036	≥10	pass	
	100 mV	99.426	100.006	100.577	0.038	≥10	pass	
Channel 2	12 mV	11.931	12.002	12.069	0.013	5.5	pass	
	8.00 V	7.9541	8.0013	8.0462	0.0022	≥10	pass	
	5.000 V	4.9713	5.0007	5.0289	0.0014	≥10	pass	
	3.000 V	2.9828	3.0019	3.0173	0.0011	≥10	pass	
	1.500 V	1.49139	1.50082	1.50866	0.00054	≥10	pass	
	1.000 V	0.99426	1.00007	1.00577	0.00036	≥10	pass	
Channel 1	100 mV	99.426	100.008	100.577	0.038	≥10	pass	
	12 mV	11.931	12.002	12.069	0.013	5.5	pass	
	[3] Sine Flatness, 1.15 Vrms (dB) - note 3							
	Channel 1	10 Hz	-0.0100	-0.0001	0.0100	0.0018	5.6	pass
		20 Hz	-0.0100	-0.0002	0.0100	0.0017	5.9	pass
		20 kHz	-0.0100	-0.0050	0.0100	0.0019	5.3	pass
50 kHz		-0.0300	0.0025	0.0300	0.0040	7.5	pass	
80 kHz		-0.1000	0.0026	0.1000	0.0094	≥10	pass	
Channel 2	10 Hz	-0.0100	-0.0003	0.0100	0.0018	5.6	pass	
	20 Hz	-0.0100	-0.0002	0.0100	0.0017	5.9	pass	
	20 kHz	-0.0100	-0.0033	0.0100	0.0019	5.3	pass	
	50 kHz	-0.0300	0.0039	0.0300	0.0040	7.5	pass	
	80 kHz	-0.1000	-0.0031	0.1000	0.0094	≥10	pass	
[4] DC Offset (mV), Unbal - note 5								
Channel 1	8.0 V	-20.10	-2.18	20.10	0.40	≥10	pass	
	1.000 V	-2.600	-0.555	2.600	0.046	≥10	pass	
	100 mV	-0.350	-0.050	0.350	0.024	≥10	pass	
	10 mV	-0.125	-0.050	0.125	0.024	5.2	pass	
Channel 2	8.0 V	-20.10	-4.11	20.10	0.40	≥10	pass	
	1.000 V	-2.600	-0.788	2.600	0.046	≥10	pass	
	100 mV	-0.350	-0.073	0.350	0.024	≥10	pass	
	10 mV	-0.125	-0.074	0.125	0.024	5.2	pass	
[5] Source Resistance Accuracy (Ω) - note 4								
Channel 1 Unbalanced	50 Ω	49.250	50.132	50.750	0.060	≥10	pass	
	600 Ω	594.00	599.61	606.00	0.28	≥10	pass	
Channel 1 Balanced	100 Ω	99.00	100.18	101.00	0.10	≥10	pass	
	600 Ω	594.00	599.43	606.00	0.30	≥10	pass	
Channel 2 Unbalanced	50 Ω	49.250	50.072	50.750	0.060	≥10	pass	
	600 Ω	594.00	599.36	606.00	0.28	≥10	pass	
Channel 2 Balanced	100 Ω	99.00	100.07	101.00	0.10	≥10	pass	
	600 Ω	594.00	599.45	606.00	0.30	≥10	pass	

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER CHARACTERISTICS							
[6] Input Termination Accuracy (Ω) - note 4							
Channel 1	600 Ω	594.00	600.28	606.00	0.30	≥ 10	pass
Channel 2	600 Ω	594.00	600.43	606.00	0.30	≥ 10	pass
[7] DC Measurement Accuracy (Volts, mVolts) - note 7							
Channel 1	+120 V	119.080	119.981	120.920	0.072	≥ 10	pass
	+80 V	79.360	79.993	80.640	0.048	≥ 10	pass
	+25 V	24.800	24.997	25.200	0.015	≥ 10	pass
	+8.0 V	7.9360	7.9990	8.0640	0.0048	≥ 10	pass
	+2.5 V	2.4800	2.5000	2.5200	0.0015	≥ 10	pass
	+800 mV	793.40	799.93	806.60	0.48	≥ 10	pass
	+250 mV	247.25	249.85	252.75	0.15	≥ 10	pass
	0 mV	-1.000	-0.114	1.000	0.046	≥ 10	pass
	-250 mV	-252.75	-250.08	-247.25	0.15	≥ 10	pass
	-800 mV	-806.60	-800.03	-793.40	0.48	≥ 10	pass
	-2.5 V	-2.5200	-2.4998	-2.4800	0.0015	≥ 10	pass
	-8 V	-8.0640	-7.9979	-7.9360	0.0048	≥ 10	pass
	-25 V	-25.200	-24.994	-24.800	0.015	≥ 10	pass
	-80 V	-80.640	-79.982	-79.360	0.048	≥ 10	pass
Channel 2	+120 V	119.080	119.968	120.920	0.072	≥ 10	pass
	+80 V	79.360	79.977	80.640	0.048	≥ 10	pass
	+25 V	24.800	24.992	25.200	0.015	≥ 10	pass
	+8.0 V	7.9360	7.9974	8.0640	0.0048	≥ 10	pass
	+2.5 V	2.4800	2.4994	2.5200	0.0015	≥ 10	pass
	+800 mV	793.40	799.88	806.60	0.48	≥ 10	pass
	+250 mV	247.25	249.98	252.75	0.15	≥ 10	pass
	0 mV	-1.000	0.054	1.000	0.046	≥ 10	pass
	-250 mV	-252.75	-249.87	-247.25	0.15	≥ 10	pass
	-800 mV	-806.60	-799.84	-793.40	0.48	≥ 10	pass
	-2.5 V	-2.5200	-2.4997	-2.4800	0.0015	≥ 10	pass
	-8 V	-8.0640	-7.9988	-7.9360	0.0048	≥ 10	pass
	-25 V	-25.200	-24.997	-24.800	0.015	≥ 10	pass
	-80 V	-80.640	-79.993	-79.360	0.048	≥ 10	pass
-120V	-120.920	-119.982	-119.080	0.072	≥ 10	pass	
[8] Input Common Mode Rejection (mV) - non-accredited							
Channel 1 (5V CM signal)	2.5V range, 200 Hz	0	0.019	0.500	0.070	na	pass
	2.5V range, 5 kHz	0	0.013	0.500	0.082	na	pass
	2.5V range, 20 kHz	0	0.03	1.58	0.18	na	pass
	8V range, 20 kHz	0	0.59	15.81	0.22	na	pass
	80V range, 20 kHz	0	3.58	28.12	0.27	na	pass
Channel 2 (5V CM signal)	2.5V range, 200 Hz	0	0.030	0.500	0.070	na	pass
	2.5V range, 5 kHz	0	0.068	0.500	0.082	na	pass
	2.5V range, 20 kHz	0	0.25	1.58	0.18	na	pass
	8V range, 20 kHz	0	0.66	15.81	0.22	na	pass
	80V range, 20 kHz	0	3.74	28.12	0.27	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[9] Level Meter AC Accuracy, 1 kHz (Volts, mVolts) - note 6							
Channel 1	85 V	84.512	84.986	85.491	0.048	≥10	pass
	20 V	19.885	19.996	20.115	0.011	≥10	pass
	5 V	4.9713	4.9990	5.0289	0.0029	≥10	pass
	2 V	1.9885	2.0000	2.0115	0.0011	≥10	pass
	500 mV	497.13	500.00	502.89	0.29	≥10	pass
	200 mV	198.85	199.98	201.15	0.11	≥10	pass
	5 mV	4.9713	5.0000	5.0289	0.0031	9.7	pass
Channel 2	85 V	84.512	84.984	85.491	0.048	≥10	pass
	20 V	19.885	19.995	20.115	0.011	≥10	pass
	5 V	4.9713	4.9987	5.0289	0.0029	≥10	pass
	2 V	1.9885	1.9997	2.0115	0.0011	≥10	pass
	500 mV	497.13	499.93	502.89	0.29	≥10	pass
	200 mV	198.85	199.95	201.15	0.11	≥10	pass
	5 mV	4.9713	4.9990	5.0289	0.0031	9.7	pass
[10] Level Meter AC Flatness, 1.15 Vrms (dB) - note 6							
Channel 1	10 Hz	-0.0100	-0.0014	0.0100	0.0030	3.3	pass
	20 Hz	-0.0100	-0.0005	0.0100	0.0029	3.4	pass
	20 kHz	-0.0100	0.0014	0.0100	0.0024	4.2	pass
	50 kHz	-0.0300	0.0040	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0226	0.1000	0.0061	≥10	pass
Channel 2	10 Hz	-0.0100	-0.0013	0.0100	0.0030	3.3	pass
	20 Hz	-0.0100	-0.0004	0.0100	0.0029	3.4	pass
	20 kHz	-0.0100	0.0020	0.0100	0.0024	4.2	pass
	50 kHz	-0.0300	0.0071	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0150	0.1000	0.0061	≥10	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[11] Phase Measurement Offset (Deg), DC Coupled - non-accredited, self-test							
Ch1 - Ch 2	50Hz	-0.200	0.000	0.200	0.001	na	pass
	200Hz	-0.200	0.000	0.200	0.001	na	pass
	5kHz	-0.200	0.004	0.200	0.004	na	pass
	20kHz	-0.800	0.016	0.800	0.016	na	pass
	50kHz	-2.000	0.037	2.000	0.040	na	pass
[12] Frequency Measurement Accuracy (uHz/Hz) - note 1							
	10 kHz	-3.00	-0.38	3.00	0.63	4.8	pass
[13] Input Residual Crosstalk at 20 kHz (dB) - non-accredited							
Ch2 into Ch 1		-999	-129.1	-120.0	4.0	na	pass
Ch 1 into Ch 2		-999	-128.4	-120.0	4.0	na	pass

CONFIDENTIAL

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
NON-ACCREDITED CHARACTERISTICS							
[14] Sine THD+N (dB) - non-accredited, self-test							
Channel 1	20 Hz, 16V, 20k BW	-999	-108.9	-102.0	1.0	na	pass
	Balanced	1 kHz, 16V, 20k BW	-999	-108.0	-102.0	1.0	na
Unbalanced	5 kHz, 16V, 20k BW	-999	-107.3	-102.0	1.0	na	pass
	20 kHz, 16V, 20k BW	-999	-108.1	-102.0	1.5	na	pass
	20 Hz, 2V, 20k BW	-999	-110.1	-102.0	1.0	na	pass
	1 kHz, 2V, 20k BW	-999	-109.5	-102.0	1.0	na	pass
	20 kHz, 2V, 20k BW	-999	-107.9	-102.0	1.5	na	pass
	Channel 2	20 Hz, 16V, 20k BW	-999	-108.6	-102.0	1.0	na
Balanced	1 kHz, 16V, 20k BW	-999	-108.2	-102.0	1.0	na	pass
	5 kHz, 16V, 20k BW	-999	-107.8	-102.0	1.0	na	pass
Unbalanced	20 kHz, 16V, 20k BW	-999	-109.5	-102.0	1.5	na	pass
	20 Hz, 2V, 20k BW	-999	-110.2	-102.0	1.0	na	pass
	1 kHz, 2V, 20k BW	-999	-109.7	-102.0	1.0	na	pass
	20 kHz, 2V, 20k BW	-999	-110.1	-102.0	1.5	na	pass
[15] Residual Crosstalk (dB), Output Related, 20kHz - non-accredited, self-test							
Unbalanced, 8V, 20 kHz.	2 into 1	-999	-124.3	-120.0	2.0	na	pass
	1 into 2	-999	-123.1	-120.0	2.0	na	pass
Balanced, 16V, 20 kHz.	2 into 1	-999	-158.1	-120.0	4.0	na	pass
	1 into 2	-999	-151.3	-120.0	4.0	na	pass
[16] Residual Noise (uVolts) - non-accredited, self-test							
Balanced Input, inputs shorted.	Ch 1, 20 kHz BW	0	1.05	1.40	0.06	na	pass
	Ch 2, 20 kHz BW	0	1.06	1.40	0.06	na	pass
[17] Residual SMPTE IMD (%), 4:1, 60Hz:7kHz - non-accredited, self-test							
Balanced, 16 Vrms.	Ch 1	0%	0.00102%	0.00250%	0.00030%	na	pass
	Ch 2	0%	0.00088%	0.00250%	0.00030%	na	pass
[18] Residual MOD IMD (%), 4:1, 60Hz:7kHz - non-accredited, self-test							
Balanced, 16 Vrms.	Ch 1	0%	0.00086%	0.00250%	0.00020%	na	pass
	Ch 2	0%	0.00082%	0.00250%	0.00020%	na	pass
[19] Residual DFD IMD (%), mean 19.5kHz, diff 1kHz - non-accredited, self-test							
Balanced, 16 Vrms.	Ch 1	0%	0.00016%	0.00100%	0.00020%	na	pass
	Ch 2	0%	0.00026%	0.00100%	0.00020%	na	pass
[20] Digital Output Amplitude Accuracy (Volts) - non-accredited, oscilloscope referenced							
Unbal, consumer	0.5 Vpp	0.450	0.487	0.550	0.009	na	pass
Unbal, professional	1.0 Vpp	0.900	0.980	1.100	0.016	na	pass
Balanced	5.0 Vpp	4.500	5.027	5.500	0.077	na	pass

END OF REPORT