



Calibration Services
 9290 SW Nimbus Ave
 Beaverton, Oregon 97008
 USA



Calibration Cert #2527.01

Calibration Report

Accredited Calibration

Report Number: XXXXXXXXXX-XXXXXX-X

Model: APx525
Serial Number: XXXXXXXXXX

Data Type: AS SHIPPED, NEW
Date of Cal: 30-Oct-2025

Program:
 APxCalibration.exe 24.01

Internal Module Status and Data			
AP Name	Description	Serial No.	Revision
BRIO	Backplane module	XXXXXX	600
BSL3	Clock module	XXXXXX	202
BMEG	Analog Output module	XXXXXX	408
BOTA	Analog Input module	XXXXXX	403
BAES	DIO module	XXXXXX	402
BSAT	Option AG52	XXXXXX	401
BROT	Option BW52	XXXXXX	400

Explanatory notes to the last three columns of the calibration report

"MU" - The column labeled "MU" lists the expanded measurement uncertainties derived from equipment specifications, repeatability data, and other significant sources. These are stated at a minimum confidence of 95% using a coverage factor k=2 (except as appropriate) following the recommendations in ISO/IEC 98-3 *Guide to the expression of uncertainty in measurement (GUM:1995)*, BIPM JCGM 100:2008, and NIST Technical Note 1297.

"TUR" - The column labeled "TUR" lists the test uncertainty ratio calculated by dividing the lesser of the lower and upper reading tolerances by the 95% expanded measurement uncertainty. An entry of "na" indicates [1] the specified limits are one-sided, or [2] the performance characteristic is not accredited.

"Result" - The column labeled "Result" lists color-coded assessments that the observed characteristic is within its specified limits of performance. There are three possible indications:

pass -- The *READING* is within the specified upper and lower limits reduced by guard-bands equal to the 95% expanded measurement uncertainty. The probability or risk of false acceptance is very low, typically <0.2%.

uncertain -- The *READING* is within the specified upper and lower limits, but it is close to one of the limits by an amount that is less than the 95% expanded measurement uncertainty. The probability or risk of false acceptance is elevated.

>> FAIL << -- The *READING* is outside of the specified limit range.

Accredited measurements listed in the following pages correlate to Audio Precision's Scope of Accreditation as noted:

- note 1 - Frequency Measurement
- note 2 - AC Voltage Measurement
- note 3 - AC Flatness Measurement
- note 4 - Resistance Measurement
- note 5 - DC Voltage Measurement
- note 6 - AC Voltage Source and AC Flatness Source for testing AC Measurement Equipment
- note 7 - DC Voltage Source for testing DC Measurement Equipment

This report is valid only when accompanied by a signed Certificate of Calibration.

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result	
ANALOG GENERATOR CHARACTERISTICS								
[1] Sine Frequency Accuracy (Hz) - note 1								
	10 kHz	9999.9700	10000.0016	10000.0300	0.0054	5.6	pass	
[2] Sine Amplitude Accuracy, 1 kHz (Volts, mVolts) - note 2								
Channel 1	13.33 V	13.2840	13.3302	13.3761	0.0066	7.1	pass	
	5.000 V	4.9828	5.0034	5.0173	0.0014	≥10	pass	
	3.000 V	2.9897	3.0010	3.0104	0.0011	9.6	pass	
	1.500 V	1.49483	1.50045	1.50519	0.00054	9.7	pass	
	1.000 V	0.99655	1.00004	1.00346	0.00036	9.7	pass	
	100 mV	99.655	99.995	100.346	0.038	9.2	pass	
Channel 2	12 mV	11.959	12.003	12.042	0.013	3.2	pass	
	13.33 V	13.2840	13.3316	13.3761	0.0066	7.1	pass	
	5.000 V	4.9828	5.0043	5.0173	0.0014	≥10	pass	
	3.000 V	2.9897	3.0015	3.0104	0.0011	9.6	pass	
	1.500 V	1.49483	1.50072	1.50519	0.00054	9.7	pass	
	1.000 V	0.99655	1.00019	1.00346	0.00036	9.7	pass	
Channel 1	100 mV	99.655	100.006	100.346	0.038	9.2	pass	
	12 mV	11.959	12.005	12.042	0.013	3.2	pass	
	[3] Sine Flatness, 1.15 Vrms (dB) - note 3							
	Channel 1	10 Hz	-0.0080	-0.0002	0.0080	0.0018	4.4	pass
		20 Hz	-0.0080	-0.0003	0.0080	0.0017	4.7	pass
		20 kHz	-0.0080	-0.0022	0.0080	0.0019	4.2	pass
50 kHz		-0.0300	0.0009	0.0300	0.0040	7.5	pass	
80 kHz		-0.1000	0.0265	0.1000	0.0092	≥10	pass	
Channel 2	10 Hz	-0.0080	-0.0001	0.0080	0.0018	4.4	pass	
	20 Hz	-0.0080	-0.0002	0.0080	0.0017	4.7	pass	
	20 kHz	-0.0080	-0.0020	0.0080	0.0019	4.2	pass	
	50 kHz	-0.0300	0.0017	0.0300	0.0040	7.5	pass	
	80 kHz	-0.1000	0.0313	0.1000	0.0092	≥10	pass	
[4] DC Offset (mV), Unbal - note 5								
Channel 1	10.0 V	-25.10	4.06	25.10	0.40	≥10	pass	
	1.000 V	-2.600	0.418	2.600	0.046	≥10	pass	
	100 mV	-0.350	0.050	0.350	0.024	≥10	pass	
	10 mV	-0.125	0.020	0.125	0.024	5.2	pass	
Channel 2	10.0 V	-25.10	0.14	25.10	0.40	≥10	pass	
	1.000 V	-2.600	0.026	2.600	0.046	≥10	pass	
	100 mV	-0.350	0.011	0.350	0.024	≥10	pass	
	10 mV	-0.125	0.020	0.125	0.024	5.2	pass	
[5] Source Resistance Accuracy (Ω) - note 4								
Channel 1 Unbalanced	20 Ω	19.600	20.028	20.400	0.048	8.3	pass	
	50 Ω	49.250	50.077	50.750	0.060	≥10	pass	
	75 Ω	74.100	74.959	75.900	0.070	≥10	pass	
	100 Ω	99.000	99.841	101.000	0.080	≥10	pass	
	600 Ω	594.00	599.90	606.00	0.28	≥10	pass	
Channel 1 Balanced	40 Ω	39.400	40.013	40.600	0.076	7.9	pass	
	100 Ω	99.00	100.12	101.00	0.10	≥10	pass	
	150 Ω	148.50	149.89	151.50	0.12	≥10	pass	
	200 Ω	198.00	199.66	202.00	0.14	≥10	pass	
	600 Ω	594.00	599.76	606.00	0.30	≥10	pass	
Channel 2 Unbalanced	20 Ω	19.600	20.040	20.400	0.048	8.3	pass	
	50 Ω	49.250	50.090	50.750	0.060	≥10	pass	
	75 Ω	74.100	74.971	75.900	0.070	≥10	pass	
	100 Ω	99.000	99.858	101.000	0.080	≥10	pass	
	600 Ω	594.00	599.87	606.00	0.28	≥10	pass	
Channel 2 Balanced	40 Ω	39.400	40.034	40.600	0.076	7.9	pass	
	100 Ω	99.00	100.13	101.00	0.10	≥10	pass	
	150 Ω	148.50	149.89	151.50	0.12	≥10	pass	
	200 Ω	198.00	199.66	202.00	0.14	≥10	pass	
	600 Ω	594.00	599.57	606.00	0.30	≥10	pass	

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER CHARACTERISTICS							
[6] Input Termination Accuracy (Ω) - note 4							
Channel 1	300 Ω	297.00	300.92	303.00	0.18	≥ 10	pass
	600 Ω	594.00	601.23	606.00	0.30	≥ 10	pass
Channel 2	300 Ω	297.00	301.13	303.00	0.18	≥ 10	pass
	600 Ω	594.00	601.73	606.00	0.30	≥ 10	pass
[7] DC Measurement Accuracy (Volts, mVolts) - note 7							
Channel 1	+160 V	158.720	159.975	161.280	0.096	≥ 10	pass
	+100 V	99.200	99.983	100.800	0.060	≥ 10	pass
	+30 V	29.758	29.994	30.242	0.018	≥ 10	pass
	+10 V	9.9200	9.9981	10.0800	0.0060	≥ 10	pass
	+3 V	2.9758	2.9994	3.0242	0.0018	≥ 10	pass
	+1 V	0.99200	0.99982	1.00800	0.00060	≥ 10	pass
	+300 mV	297.10	299.97	302.90	0.18	≥ 10	pass
	0 mV	-0.800	0.051	0.800	0.046	≥ 10	pass
	-300 mV	-302.90	-299.87	-297.10	0.18	≥ 10	pass
	-1 V	-1.00800	-0.99977	-0.99200	0.00060	≥ 10	pass
	-3 V	-3.0242	-2.9994	-2.9758	0.0018	≥ 10	pass
	-10 V	-10.0800	-9.9979	-9.9200	0.0060	≥ 10	pass
	-30 V	-30.242	-29.995	-29.758	0.018	≥ 10	pass
	-100V	-100.800	-99.981	-99.200	0.060	≥ 10	pass
-160 V	-161.280	-159.969	-158.720	0.096	≥ 10	pass	
Channel 2	+160 V	158.720	159.948	161.280	0.096	≥ 10	pass
	+100 V	99.200	99.969	100.800	0.060	≥ 10	pass
	+30 V	29.758	29.991	30.242	0.018	≥ 10	pass
	+10 V	9.9200	9.9969	10.0800	0.0060	≥ 10	pass
	+3 V	2.9758	2.9990	3.0242	0.0018	≥ 10	pass
	+1 V	0.99200	0.99971	1.00800	0.00060	≥ 10	pass
	+300 mV	297.10	299.93	302.90	0.18	≥ 10	pass
	0 mV	-0.800	0.056	0.800	0.046	≥ 10	pass
	-300 mV	-302.90	-299.82	-297.10	0.18	≥ 10	pass
	-1 V	-1.00800	-0.99960	-0.99200	0.00060	≥ 10	pass
	-3 V	-3.0242	-2.9990	-2.9758	0.0018	≥ 10	pass
	-10 V	-10.0800	-9.9972	-9.9200	0.0060	≥ 10	pass
	-30 V	-30.242	-29.991	-29.758	0.018	≥ 10	pass
	-100V	-100.800	-99.974	-99.200	0.060	≥ 10	pass
-160 V	-161.280	-159.963	-158.720	0.096	≥ 10	pass	
[8] Input Common Mode Rejection (mV) - non-accredited							
Channel 1 (5V CM signal)	3.2V range, 200 Hz	0	0.026	0.500	0.070	na	pass
	3.2V range, 5 kHz	0	0.064	0.500	0.082	na	pass
	3.2V range, 20 kHz	0	0.25	1.26	0.18	na	pass
	10V range, 20 kHz	0	1.39	15.81	0.23	na	pass
	100V range, 20 kHz	0	3.01	28.12	0.27	na	pass
Channel 2 (5V CM signal)	3.2V range, 200 Hz	0	0.022	0.500	0.070	na	pass
	3.2V range, 5 kHz	0	0.103	0.500	0.082	na	pass
	3.2V range, 20 kHz	0	0.41	1.26	0.18	na	pass
	10V range, 20 kHz	0	0.49	15.81	0.23	na	pass
	100V range, 20 kHz	0	5.64	28.12	0.27	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[9] Level Meter AC Accuracy, 1 kHz (Volts, mVolts) - note 6							
Channel 1	160 V	159.448	159.975	160.554	0.090	6.2	pass
	90 V	89.690	89.985	90.311	0.050	6.3	pass
	30 V	29.897	29.995	30.104	0.017	6.2	pass
	9 V	8.9690	8.9981	9.0311	0.0050	6.2	pass
	3 V	2.9897	2.9995	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89985	0.90311	0.00050	6.2	pass
	300 mV	298.97	299.94	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9991	5.0173	0.0031	5.6	pass
500 kHz BW	3 V	2.9897	2.9987	3.0104	0.0017	6.2	pass
Channel 2	160 V	159.45	159.96	160.55	0.090	6.2	pass
	90 V	89.690	89.976	90.311	0.050	6.3	pass
	30 V	29.897	29.991	30.104	0.017	6.2	pass
	9 V	8.9690	8.9973	9.0311	0.0050	6.2	pass
	3 V	2.9897	2.9992	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89974	0.90311	0.00050	6.2	pass
	300 mV	298.97	299.90	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9984	5.0173	0.0031	5.6	pass
500 kHz BW	3 V	2.9897	2.9982	3.0104	0.0017	6.2	pass
[10] Level Meter AC Flatness, 1.15 Vrms (dB) - note 6							
Channel 1	10 Hz	-0.0080	-0.0013	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0002	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0021	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0068	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0147	0.1000	0.0061	≥10	pass
	100 kHz	-0.1000	-0.0093	0.1000	0.0061	≥10	pass
	200 kHz	-0.200	-0.033	0.200	0.021	9.5	pass
500 kHz BW	100 kHz	-0.1000	-0.0093	0.1000	0.0061	≥10	pass
Channel 2	10 Hz	-0.0080	-0.0013	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0002	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0021	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0063	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0157	0.1000	0.0061	≥10	pass
	100 kHz	-0.2000	-0.0075	0.2000	0.0061	≥10	pass
	200 kHz	-0.200	-0.026	0.200	0.021	9.5	pass
500 kHz BW	200 kHz	-0.200	-0.026	0.200	0.021	9.5	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[11] Phase Measurement Offset (Deg) - non-accredited, self-test							
Ch1 - Ch 2	50Hz, AC coupled	-0.200	0.003	0.200	0.001	na	pass
	200Hz, DC coupled	-0.200	0.000	0.200	0.001	na	pass
	5kHz, DC coupled	-0.200	0.005	0.200	0.004	na	pass
	20kHz, DC coupled	-0.800	0.019	0.800	0.016	na	pass
	50kHz, DC coupled	-2.000	0.051	2.000	0.040	na	pass
[12] Frequency Measurement Accuracy (uHz/Hz) - note 1							
	10 kHz	-3.00	-0.16	3.00	0.54	5.6	pass
[13] Input Residual Crosstalk at 20 kHz (dB) - non-accredited							
Ch2 into Ch 1		-999	-149.7	-130.0	4.0	na	pass
Ch 1 into Ch 2		-999	-150.8	-130.0	4.0	na	pass
OPTION AG52 CHARACTERISTICS							
[14] Squarewave Amplitude Accuracy (Volts) - note 2							
	100 Hz, 1 Vpk	0.9886	0.9997	1.0116	0.0008	≥10	pass
[15] Squarewave Risetime (usec) - non-accredited, oscilloscope referenced							
	5 kHz, 1 Vpk	0.00	1.64	2.00	0.06	na	pass
[16] Squarewave Even Harmonic Content (dB) - non-accredited, self-test							
	20 Hz, 1 Vpk	-999	-121.0	-100.0	4.0	na	pass
	5 kHz, 1 Vpk	-999	-105.4	-100.0	2.0	na	pass
	20 kHz, 1 Vpk	-999	-94.7	-90.0	1.0	na	pass
[17] DIM Amplitude Accuracy (Volts) - note 2							
	DIM100, 1Vpk	0.7958	0.8041	0.8143	0.0006	≥10	pass
[18] Residual IMD, DIM100 (%) - non-accredited, self-test							
Balanced, 21.2Vrms	Ch 1	0.00000%	0.00075%	0.00180%	0.00020%	na	pass
	Ch 2	0.00000%	0.00081%	0.00180%	0.00020%	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result	
NON-ACCREDITED CHARACTERISTICS								
[19] Sine THD+N (dB) - non-accredited, self-test								
Channel 1 <i>Balanced</i>	20 Hz, 21.2V, 20k BW	-999	-112.1	-105.0	1.0	na	pass	
	1 kHz, 21.2V, 20k BW	-999	-111.9	-105.0	1.0	na	pass	
	5 kHz, 21.2V, 20k BW	-999	-111.5	-105.0	1.0	na	pass	
	20 kHz, 21.2V, 20k BW	-999	-112.5	-105.0	1.5	na	pass	
	<i>Unbalanced</i>	20 Hz, 2.5V, 20k BW	-999	-114.2	-105.0	1.0	na	pass
		1 kHz, 2.5V, 20k BW	-999	-114.0	-105.0	1.0	na	pass
		20 kHz, 2.5V, 20k BW	-999	-113.3	-105.0	1.5	na	pass
		1 kHz, 2.5V 40k BW	-999	-110.4	-98.0	1.0	na	pass
		10 kHz, 2.5V, 40k BW	-999	-109.1	-98.0	1.0	na	pass
		1 kHz, 2.5V 80k BW	-999	-103.1	-90.0	1.0	na	pass
		20 kHz, 2.5V, 80k BW	-999	-101.3	-90.0	1.5	na	pass
		20kHz, 2.5V, 250k BW	-999	-93.7	-85.0	1.5	na	pass
		20kHz, 2.5V, 500k BW	-999	-91.7	-82.0	1.5	na	pass
		Channel 2 <i>Balanced</i>	20 Hz, 21.2V, 20k BW	-999	-111.7	-105.0	1.0	na
1 kHz, 21.2V, 20k BW	-999		-111.7	-105.0	1.0	na	pass	
5 kHz, 21.2V, 20k BW	-999		-111.0	-105.0	1.0	na	pass	
20 kHz, 21.2V, 20k BW	-999		-112.4	-105.0	1.5	na	pass	
<i>Unbalanced</i>	20 Hz, 2.5V, 20k BW		-999	-114.5	-105.0	1.0	na	pass
	1 kHz, 2.5V, 20k BW		-999	-113.5	-105.0	1.0	na	pass
	20 kHz, 2.5V, 20k BW		-999	-113.3	-105.0	1.5	na	pass
	1 kHz, 2.5V 40k BW		-999	-110.2	-98.0	1.0	na	pass
	10 kHz, 2.5V, 40k BW		-999	-108.9	-98.0	1.0	na	pass
	1 kHz, 2.5V 80k BW		-999	-103.0	-90.0	1.0	na	pass
	20 kHz, 2.5V, 80k BW		-999	-101.3	-90.0	1.5	na	pass
	20kHz, 2.5V, 250k BW		-999	-93.9	-85.0	1.5	na	pass
	20kHz, 2.5V, 500k BW		-999	-91.8	-82.0	1.5	na	pass
	[20] Residual Crosstalk (dB), Output Related, 20kHz - non-accredited, self-test							
<i>Unbalanced, 10.6V, 20 kHz.</i>	2 into 1	-999	-152.4	-130.0	4.0	na	pass	
	1 into 2	-999	-156.2	-130.0	4.0	na	pass	
<i>Balanced, 21.2V, 20 kHz.</i>	2 into 1	-999	-159.3	-130.0	4.0	na	pass	
	1 into 2	-999	-163.4	-130.0	4.0	na	pass	
[21] Residual Noise (uVolts) - non-accredited, self-test								
<i>Balanced Input, inputs shorted.</i>	Ch 1, 20 kHz BW	0	0.97	1.30	0.06	na	pass	
	Ch 2, 20 kHz BW	0	0.97	1.30	0.06	na	pass	
	Ch 1, 500 kHz BW	0	5.74	11.40	0.20	na	pass	
	Ch 2, 500 kHz BW	0	5.71	11.40	0.20	na	pass	
[22] Residual SMPTE IMD (%), 4:1, 60Hz:7kHz - non-accredited, self-test								
<i>Balanced, 21.2 Vrms.</i>	Ch 1	0%	0.00048%	0.00180%	0.00030%	na	pass	
	Ch 2	0%	0.00059%	0.00180%	0.00030%	na	pass	
[23] Residual MOD IMD (%), 4:1, 60Hz:7kHz - non-accredited, self-test								
<i>Balanced, 21.2 Vrms.</i>	Ch 1	0%	0.00044%	0.00180%	0.00020%	na	pass	
	Ch 2	0%	0.00058%	0.00180%	0.00020%	na	pass	
[24] Residual DFD IMD (%), mean 19.5kHz, diff 1kHz - non-accredited, self-test								
<i>Balanced, 21.2 Vrms.</i>	Ch 1	0%	0.00010%	0.00050%	0.00012%	na	pass	
	Ch 2	0%	0.00011%	0.00050%	0.00012%	na	pass	
[25] Digital Output Amplitude Accuracy (Volts) - non-accredited								
<i>Unbal, consumer</i>	0.5 Vpp	0.450	0.497	0.550	0.009	na	pass	
<i>Unbal, professional</i>	1.0 Vpp	0.900	0.984	1.100	0.014	na	pass	
<i>Balanced</i>	5.0 Vpp	4.500	5.058	5.500	0.077	na	pass	

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ADVANCED DIGITAL I/O CHARACTERISTICS				>> Option Not Installed			
[26] Output Voltage Accuracy (Volts) - non-accredited, oscilloscope referenced							
[27] Input Voltage Measurement Accuracy (Volts) - non-accredited, oscilloscope referenced							
[28] Variable Rise Time Accuracy (nsec) - non-accredited, oscilloscope referenced							
[29] Normal Mode Noise Amplitude (Volts) - non-accredited, oscilloscope referenced							
[30] Common Mode Sine Amplitude (Volts) - note 2							
[31] Jitter Amplitude Accuracy at 500 Hz (nsec) - non-accredited, self-test							
[32] Jitter Flatness (dB) - non-accredited, self-test							
[33] Residual Jitter, 700 Hz - 100 kHz BW (psec, pk) - non-accredited, self-test							

END OF REPORT