



Calibration Services  
 9290 SW Nimbus Ave  
 Beaverton, Oregon 97008  
 USA



## Calibration Report

### Accredited Calibration

### Report Number: 101305555-260415-1

**Model:** APx555  
**Serial Number:** XXXXXXXXX

**Data Type:** AS SHIPPED, NEW  
**Date of Cal:** XXXXXX

**Program:**  
 APxCalibration.exe 24.01

Internal Module Status and Data			
AP Name	Description	Serial No.	Revision
BRIO	Backplane module	XXXXX	600
BSL4	Advanced Master Clock	XXXXX	201
B4AO	Analog Output module	XXXXX	401
B4SG	Analog Oscillator module	XXXXX	304
B4AI	Analog Input module	XXXXX	400
B4BR	Analog Bandreject module	XXXXX	400
BADI	Advanced DIO module	XXXXX	103

Software Options	
ADCTest	Installed

#### Explanatory notes to the last three columns of the calibration report

**"MU"** - The column labeled "MU" lists the expanded measurement uncertainties derived from published specifications of the calibration equipment, resolution factors, statistical data of repeatability and/or other information. These are based on a coverage factor of 2 (k=2) corresponding to a confidence level of 95% using the recommendations in the Guide to the Expression of Uncertainty in Measurement, 1993, ISO Geneva; and NIST Technical Note 1297 (1994).

**"TUR"** - The column labeled "TUR" lists the test uncertainty ratio calculated by dividing the lesser of the lower and upper reading tolerances by the 95% expanded measurement uncertainty. An entry of "na" indicates [1] the specified limits are one-sided, or [2] the performance characteristic is a self-test or system specification that is not independently certifiable.

**"Result"** - The column labeled "Result" lists color-coded assessments that the observed characteristic is within its specified limits of performance. There are three possible indications:

**pass** -- The *READING* is within the specified upper and lower limits reduced by guard-bands equal to the 95% expanded measurement uncertainty. The probability or risk of false acceptance is very low, typically <0.2%.

**uncertain** -- The *READING* is within the specified upper and lower limits, but it is close to one of the limits by an amount that is less than the 95% expanded measurement uncertainty. The probability or risk of false acceptance is elevated.

**>> FAIL <<** -- The *READING* is outside of the specified limit range.

Accredited measurements listed in the following pages correlate to Audio Precision's Scope of Accreditation as noted:

- note 1 - Frequency Measurement
- note 2 - AC Voltage Measurement
- note 3 - AC Flatness Measurement
- note 4 - Resistance Measurement
- note 5 - DC Voltage Measurement
- note 6 - AC Voltage Source and AC Flatness Source for testing AC Measurement Equipment
- note 7 - DC Voltage Source for testing DC Measurement Equipment

This report is valid only when accompanied by a signed Certificate of Calibration.

Specifications: NP0020.00030 r003

File name: APx555CalReport.xlsx

File version: REV4.34-02FEB2026-LB

NP0020.00041 r002

Uncertainty Calcs:

202602

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
<b>ANALOG GENERATOR CHARACTERISTICS</b>							
<b>[1] DAC Sine Frequency Accuracy (Hz) - note 1</b>							
	10 kHz	9999.9700	<b>10000.0009</b>	10000.0300	0.0054	5.6	pass
<b>[2] Analog Sine Frequency Accuracy (Hz) - note 1</b>							
Precision Tune	5 Hz	4.9825	<b>5.0008</b>	5.0175	0.0020	≥10	pass
	10 Hz	9.9650	<b>9.9981</b>	10.0350	0.0020	≥10	pass
	100 Hz	99.650	<b>100.002</b>	100.350	0.020	≥10	pass
	1 kHz	996.50	<b>999.95</b>	1003.50	0.20	≥10	pass
	10 kHz	9965.0	<b>9999.5</b>	10035.0	2.0	≥10	pass
	100 kHz	99650	<b>99999</b>	100350	30	≥10	pass
	200 kHz	199000	<b>200106</b>	201000	60	≥10	pass
	5 Hz	4.99970	<b>4.99998</b>	5.00030	0.00005	6.0	pass
	10 Hz	9.99970	<b>10.00001</b>	10.00030	0.00008	3.8	pass
	100 Hz	99.9970	<b>100.0000</b>	100.0030	0.0004	7.5	pass
	1 kHz	999.970	<b>1000.000</b>	1000.030	0.004	7.5	pass
	10 kHz	9999.70	<b>10000.00</b>	10000.30	0.04	7.5	pass
	100 kHz	99997.0	<b>100000.0</b>	100003.0	0.4	7.5	pass
	200 kHz	199994.0	<b>199999.5</b>	200006.0	0.8	7.5	pass
<b>[3] DAC Sine Amplitude Accuracy, 1 kHz (Volts, mVolts) - note 2</b>							
Channel 1	13.33 V	13.2840	<b>13.3312</b>	13.3761	0.0066	7.0	pass
	10.00V	9.9655	<b>9.9998</b>	10.0346	0.0028	≥10	pass
	5.000V	4.9828	<b>5.0001</b>	5.0173	0.0014	≥10	pass
	2.000V	1.99310	<b>2.00017</b>	2.00692	0.00072	9.6	pass
	1.500V	1.49483	<b>1.50015</b>	1.50519	0.00054	9.6	pass
	1.000V	0.99655	<b>1.00023</b>	1.00346	0.00028	≥10	pass
	500 mV	498.28	<b>500.03</b>	501.73	0.14	≥10	pass
	300 mV	298.97	<b>300.02</b>	301.04	0.11	9.4	pass
	200 mV	199.310	<b>200.042</b>	200.692	0.085	8.1	pass
	100 mV	99.655	<b>100.015</b>	100.346	0.038	9.1	pass
	70 mV	69.759	<b>70.019</b>	70.242	0.032	7.6	pass
	50 mV	49.828	<b>49.997</b>	50.173	0.023	7.5	pass
	20 mV	19.931	<b>19.998</b>	20.069	0.014	4.9	pass
	12 mV	11.959	<b>12.002</b>	12.042	0.013	3.2	pass
Channel 2	13.33 V	13.2840	<b>13.3298</b>	13.3761	0.0066	7.0	pass
	10.00V	9.9655	<b>9.9997</b>	10.0346	0.0028	≥10	pass
	5.000V	4.9828	<b>4.9999</b>	5.0173	0.0014	≥10	pass
	2.000V	1.99310	<b>2.00006</b>	2.00692	0.00072	9.6	pass
	1.500V	1.49483	<b>1.50008</b>	1.50519	0.00054	9.6	pass
	1.000V	0.99655	<b>1.00017</b>	1.00346	0.00028	≥10	pass
	500 mV	498.28	<b>500.09</b>	501.73	0.14	≥10	pass
	300 mV	298.97	<b>300.04</b>	301.04	0.11	9.4	pass
	200 mV	199.310	<b>200.017</b>	200.692	0.085	8.1	pass
	100 mV	99.655	<b>100.019</b>	100.346	0.038	9.1	pass
	70 mV	69.759	<b>70.013</b>	70.242	0.032	7.6	pass
	50 mV	49.828	<b>49.993</b>	50.173	0.023	7.5	pass
	20 mV	19.931	<b>19.999</b>	20.069	0.014	4.9	pass
	12 mV	11.959	<b>12.002</b>	12.042	0.013	3.2	pass
<b>[4] Analog Sine Amplitude Accuracy, 1 kHz (Volts) - note 2</b>							
Channel 1	1 V	0.99655	<b>1.00036</b>	1.00346	0.00028	≥10	pass
Channel 2	1 V	0.99655	<b>1.00030</b>	1.00346	0.00028	≥10	pass
<b>[5] Sine Burst Amplitude Accuracy, 1 kHz (Volts) - note 2</b>							
	1 V	0.99426	<b>1.00019</b>	1.00577	0.00028	≥10	pass
<b>[6] Sine Burst Ratio Accuracy, 1 kHz (dB) - note 2</b>							
	0 dB	-0.150	<b>-0.029</b>	0.150	0.017	8.8	pass
	-10 dB	-10.150	<b>-10.024</b>	-9.850	0.020	7.5	pass
	-20 dB	-20.150	<b>-20.035</b>	-19.850	0.021	7.1	pass
	-30 dB	-30.150	<b>-30.029</b>	-29.850	0.020	7.5	pass
	-40 dB	-40.150	<b>-40.035</b>	-39.850	0.021	7.1	pass
	-50 dB	-50.150	<b>-50.030</b>	-49.850	0.040	3.8	pass
	-60 dB	-60.150	<b>-60.030</b>	-59.850	0.040	3.8	pass
<b>[7] SMPTE IMD Amplitude Accuracy, 60 Hz + 7 kHz (Volts) - note 2</b>							
	10:1 ratio	0.9063	<b>0.9136</b>	0.9210	0.0006	≥10	pass
	4:1 ratio	0.8180	<b>0.8246</b>	0.8313	0.0006	≥10	pass
	1:1 ratio	0.7014	<b>0.7072</b>	0.7128	0.0005	≥10	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
<b>ANALOG GENERATOR, continued</b>							
<b>[8] DFD IMD Amplitude Accuracy, 19 kHz + 20 kHz (Volts) - note 2</b>							
	1 V	0.7014	<b>0.7070</b>	0.7128	0.0005	≥10	pass
<b>[9] DIM IMD Amplitude Accuracy (Volts) - note 2</b>							
DIM-100	1 V	1.1282	<b>1.1384</b>	1.1519	0.0024	5.0	pass
DIM-30	1 V	1.0995	<b>1.1118</b>	1.1226	0.0024	4.8	pass
<b>[10] Square-Wave Amplitude Accuracy, 100 Hz (Volts) - note 2</b>							
	1 Vrms	1.3996	<b>1.4142</b>	1.4289	0.0035	4.2	pass
<b>[11] DAC Sine Flatness, 1.15 Vrms (dB) - note 3</b>							
Channel 1	10 Hz	-0.0080	<b>-0.0004</b>	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	<b>-0.0003</b>	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	<b>-0.0025</b>	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	<b>-0.0021</b>	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	<b>-0.0405</b>	0.1000	0.0092	≥10	pass
Channel 2	10 Hz	-0.0080	<b>-0.0004</b>	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	<b>-0.0002</b>	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	<b>-0.0027</b>	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	<b>-0.0023</b>	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	<b>-0.0409</b>	0.1000	0.0092	≥10	pass
<b>[12] Analog Sine Flatness, 1.15 Vrms (dB) - note 3</b>							
Channel 1	5 Hz	-0.0080	<b>0.0002</b>	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	<b>0.0000</b>	0.0080	0.0017	4.7	pass
	50 Hz	-0.0080	<b>-0.0001</b>	0.0080	0.0014	5.7	pass
	10 kHz	-0.0080	<b>-0.0005</b>	0.0080	0.0019	4.2	pass
	20 kHz	-0.0080	<b>-0.0001</b>	0.0080	0.0019	4.2	pass
	50 kHz	-0.0150	<b>0.0019</b>	0.0150	0.0036	4.2	pass
	100 kHz	-0.0600	<b>0.0087</b>	0.0600	0.0081	7.4	pass
Channel 2	200 kHz	-0.150	<b>0.014</b>	0.150	0.028	5.4	pass
	5 Hz	-0.0080	<b>0.0000</b>	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	<b>0.0000</b>	0.0080	0.0017	4.7	pass
	50 Hz	-0.0080	<b>-0.0001</b>	0.0080	0.0014	5.7	pass
	10 kHz	-0.0080	<b>-0.0004</b>	0.0080	0.0019	4.2	pass
	20 kHz	-0.0080	<b>-0.0002</b>	0.0080	0.0019	4.2	pass
	50 kHz	-0.0150	<b>0.0020</b>	0.0150	0.0036	4.2	pass
100 kHz	-0.0600	<b>0.0088</b>	0.0600	0.0081	7.4	pass	
200 kHz	-0.150	<b>0.015</b>	0.150	0.028	5.4	pass	
<b>[13] DC Offset (mV), Unbal - note 5</b>							
Channel 1	10.0 V	-25.10	<b>2.01</b>	25.10	0.40	≥10	pass
	1.000 V	-2.600	<b>0.219</b>	2.600	0.40	6.5	pass
	100 mV	-0.350	<b>0.026</b>	0.350	0.024	≥10	pass
	10 mV	-0.125	<b>0.005</b>	0.125	0.024	5.2	pass
Channel 2	10.0 V	-25.10	<b>1.03</b>	25.10	0.40	≥10	pass
	1.000 V	-2.600	<b>0.200</b>	2.600	0.40	6.5	pass
	100 mV	-0.350	<b>0.023</b>	0.350	0.024	≥10	pass
	10 mV	-0.125	<b>0.004</b>	0.125	0.024	5.2	pass
<b>[14] Source Resistance Accuracy (Ω) - note 4</b>							
Ch 1, Unbal	20 Ω	19.600	<b>20.120</b>	20.400	0.048	8.3	pass
	50 Ω	49.250	<b>50.134</b>	50.750	0.060	≥10	pass
	75 Ω	74.100	<b>75.129</b>	75.900	0.070	≥10	pass
	100 Ω	99.000	<b>100.124</b>	101.000	0.080	≥10	pass
	600 Ω	597.00	<b>599.72</b>	603.00	0.28	≥10	pass
Ch 2, Unbal	20 Ω	19.600	<b>20.115</b>	20.400	0.048	8.3	pass
	50 Ω	49.250	<b>50.136</b>	50.750	0.060	≥10	pass
	75 Ω	74.100	<b>75.130</b>	75.900	0.070	≥10	pass
	100 Ω	99.000	<b>100.127</b>	101.000	0.080	≥10	pass
	600 Ω	597.00	<b>599.69</b>	603.00	0.28	≥10	pass
Ch 1, Bal	40 Ω	39.400	<b>40.270</b>	40.600	0.076	7.9	pass
	100 Ω	99.00	<b>100.30</b>	101.00	0.10	≥10	pass
	150 Ω	148.50	<b>150.29</b>	151.50	0.12	≥10	pass
	200 Ω	198.00	<b>200.28</b>	202.00	0.14	≥10	pass
Ch 2, Bal	600 Ω	597.00	<b>600.11</b>	603.00	0.30	≥10	pass
	40 Ω	39.400	<b>40.251</b>	40.600	0.076	7.9	pass
	100 Ω	99.00	<b>100.28</b>	101.00	0.10	≥10	pass
	150 Ω	148.50	<b>150.28</b>	151.50	0.12	≥10	pass
	200 Ω	198.00	<b>200.27</b>	202.00	0.14	≥10	pass
	600 Ω	597.00	<b>600.09</b>	603.00	0.30	≥10	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
<b>ANALOG GENERATOR, continued</b>							
<b>[15] DC Offset Accuracy (V), Bal ADC Test - note 5</b>				ADC Test requires a purchased software option.			
Channel 1 pos 200 mVrms	4.2 V	4.1844	<b>4.1969</b>	4.2156	0.0004	≥10	pass
	2.5 V	2.4895	<b>2.4986</b>	2.5105	0.0004	≥10	pass
	-0.4 V	-0.4042	<b>-0.3990</b>	-0.3958	0.0004	≥10	pass
Channel 1 neg 200 mVrms	4.2 V	4.1844	<b>4.1969</b>	4.2156	0.0004	≥10	pass
	2.5 V	2.4895	<b>2.4986</b>	2.5105	0.0004	≥10	pass
	-0.4 V	-0.4042	<b>-0.3991</b>	-0.3958	0.0004	≥10	pass
Channel 1 pos 100 mVrms	4.2 V	4.1844	<b>4.1988</b>	4.2156	0.0004	≥10	pass
	2.5 V	2.4895	<b>2.4999</b>	2.5105	0.0004	≥10	pass
	-0.4 V	-0.4042	<b>-0.3991</b>	-0.3958	0.0004	≥10	pass
Channel 1 neg 100 mVrms	4.2 V	4.1844	<b>4.1988</b>	4.2156	0.0004	≥10	pass
	2.5 V	2.4895	<b>2.4998</b>	2.5105	0.0004	≥10	pass
	-0.4 V	-0.4042	<b>-0.3991</b>	-0.3958	0.0004	≥10	pass
Channel 2 pos 200 mVrms	4.2 V	4.1844	<b>4.1973</b>	4.2156	0.0004	≥10	pass
	2.5 V	2.4895	<b>2.4986</b>	2.5105	0.0004	≥10	pass
	-0.4 V	-0.4042	<b>-0.3998</b>	-0.3958	0.0004	≥10	pass
Channel 2 neg 200 mVrms	4.2 V	4.1844	<b>4.1972</b>	4.2156	0.0004	≥10	pass
	2.5 V	2.4895	<b>2.4985</b>	2.5105	0.0004	≥10	pass
	-0.4 V	-0.4042	<b>-0.3999</b>	-0.3958	0.0004	≥10	pass
Channel 2 pos 100 mVrms	4.2 V	4.1844	<b>4.1991</b>	4.2156	0.0004	≥10	pass
	2.5 V	2.4895	<b>2.4998</b>	2.5105	0.0004	≥10	pass
	-0.4 V	-0.4042	<b>-0.3999</b>	-0.3958	0.0004	≥10	pass
Channel 2 neg 100 mVrms	4.2 V	4.1844	<b>4.1991</b>	4.2156	0.0004	≥10	pass
	2.5 V	2.4895	<b>2.4997</b>	2.5105	0.0004	≥10	pass
	-0.4 V	-0.4042	<b>-0.4000</b>	-0.3958	0.0004	≥10	pass

CONFIDENTIAL

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
<b>ANALOG ANALYZER CHARACTERISTICS</b>							
<b>[16] Input Termination Accuracy (<math>\Omega</math>) - note 4</b>							
Channel 1	300 $\Omega$	298.50	<b>300.67</b>	301.50	0.18	8.3	pass
	600 $\Omega$	597.00	<b>600.98</b>	603.00	0.30	$\geq 10$	pass
Channel 2	300 $\Omega$	298.50	<b>300.61</b>	301.50	0.18	8.3	pass
	600 $\Omega$	597.00	<b>600.95</b>	603.00	0.30	$\geq 10$	pass
<b>[17] DC Measurement Accuracy (Volts, mVolts) - note 7</b>							
Channel 1	+160V, 320V range	158.560	<b>159.911</b>	161.440	0.260	5.5	pass
	+160 V	158.720	<b>159.916</b>	161.280	0.260	4.9	pass
	+80 V	79.360	<b>79.957</b>	80.640	0.130	4.9	pass
	+40 V	39.680	<b>39.979</b>	40.320	0.064	5.0	pass
	+20 V	19.840	<b>19.989</b>	20.160	0.032	5.0	pass
	+10 V	9.9200	<b>9.9944</b>	10.0800	0.0160	5.0	pass
	+5 V	4.9600	<b>4.9971</b>	5.0400	0.0080	5.0	pass
	+2.5 V	2.4800	<b>2.4987</b>	2.5200	0.0040	5.0	pass
	+1.25 V	1.2400	<b>1.2494</b>	1.2600	0.0020	5.0	pass
	+620 mV	614.86	<b>619.72</b>	625.14	0.99	5.2	pass
	+310 mV	307.03	<b>309.85</b>	312.97	0.50	5.9	pass
	0 mV	-0.800	<b>0.021</b>	0.800	0.160	5.0	pass
	-310 mV	-312.97	<b>-309.81</b>	-307.03	0.50	5.9	pass
	-620 mV	-625.14	<b>-619.69</b>	-614.86	0.99	5.2	pass
	-1.25 V	-1.2600	<b>-1.2494</b>	-1.2400	0.0020	5.0	pass
	-2.5 V	-2.5200	<b>-2.4989</b>	-2.4800	0.0040	5.0	pass
	-5 V	-5.0400	<b>-4.9975</b>	-4.9600	0.0080	5.0	pass
	-10 V	-10.0800	<b>-9.9950</b>	-9.9200	0.0160	5.0	pass
	-20 V	-20.160	<b>-19.990</b>	-19.840	0.032	5.0	pass
	-40V	-40.320	<b>-39.981</b>	-39.680	0.064	5.0	pass
-80 V	-80.640	<b>-79.964</b>	-79.360	0.130	4.9	pass	
-160 V	-161.280	<b>-159.926</b>	-158.720	0.260	4.9	pass	
-160V, 320V range	-161.440	<b>-159.932</b>	-158.560	0.260	5.5	pass	
Channel 2	+160V, 320V range	158.560	<b>159.935</b>	161.440	0.260	5.5	pass
	+160 V	158.720	<b>159.919</b>	161.280	0.260	4.9	pass
	+80 V	79.360	<b>79.958</b>	80.640	0.130	4.9	pass
	+40 V	39.680	<b>39.979</b>	40.320	0.064	5.0	pass
	+20 V	19.840	<b>19.989</b>	20.160	0.032	5.0	pass
	+10 V	9.9200	<b>9.9951</b>	10.0800	0.0160	5.0	pass
	+5 V	4.9600	<b>4.9975</b>	5.0400	0.0080	5.0	pass
	+2.5 V	2.4800	<b>2.4984</b>	2.5200	0.0040	5.0	pass
	+1.25 V	1.2400	<b>1.2492</b>	1.2600	0.0020	5.0	pass
	+620 mV	614.86	<b>619.54</b>	625.14	0.99	5.2	pass
	+310 mV	307.03	<b>309.69</b>	312.97	0.50	5.9	pass
	0 mV	-0.800	<b>-0.076</b>	0.800	0.160	5.0	pass
	-310 mV	-312.97	<b>-309.84</b>	-307.03	0.50	5.9	pass
	-620 mV	-625.14	<b>-619.61</b>	-614.86	0.99	5.2	pass
	-1.25 V	-1.2600	<b>-1.2491</b>	-1.2400	0.0020	5.0	pass
	-2.5 V	-2.5200	<b>-2.4981</b>	-2.4800	0.0040	5.0	pass
	-5 V	-5.0400	<b>-4.9967</b>	-4.9600	0.0080	5.0	pass
	-10 V	-10.0800	<b>-9.9934</b>	-9.9200	0.0160	5.0	pass
	-20 V	-20.160	<b>-19.988</b>	-19.840	0.032	5.0	pass
	-40V	-40.320	<b>-39.974</b>	-39.680	0.064	5.0	pass
-80 V	-80.640	<b>-79.950</b>	-79.360	0.130	4.9	pass	
-160 V	-161.280	<b>-159.896</b>	-158.720	0.260	4.9	pass	
-160V, 320V range	-161.440	<b>-159.885</b>	-158.560	0.260	5.5	pass	

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
<b>ANALOG ANALYZER, continued</b>							
<b>[18] Input Common Mode Rejection (mV) - non-accredited</b>							
Channel 1 (5V CM signal)	2.5V range, 200 Hz	0	0.003	0.158	0.070	na	pass
	2.5V range, 5 kHz	0	0.010	0.158	0.082	na	pass
	2.5V range, 20 kHz	0	0.03	0.50	0.18	na	pass
	5V range, 20 kHz	0	0.41	2.81	0.19	na	pass
	20V range, 5 kHz	0	7.42	63.25	0.87	na	pass
	80V range, 5 kHz	0	4.08	63.25	0.87	na	pass
Channel 2 (5V CM signal)	2.5V range, 200 Hz	0	0.017	0.158	0.070	na	pass
	2.5V range, 5 kHz	0	0.021	0.158	0.082	na	pass
	2.5V range, 20 kHz	0	0.05	0.50	0.18	na	pass
	5V range, 20 kHz	0	0.30	2.81	0.19	na	pass
	20V range, 5 kHz	0	4.85	63.25	0.87	na	pass
	80V range, 5 kHz	0	6.31	63.25	0.87	na	pass
<b>[19] Level Meter AC Accuracy, 1 kHz (Volts, mVolts) - note 6</b>							
Channel 1	160 V	159.448	159.921	160.554	0.260	2.1	pass
	100 V	99.655	99.951	100.346	0.160	2.2	pass
	50 V	49.828	49.975	50.173	0.081	2.1	pass
	25 V	24.914	24.987	25.086	0.040	2.2	pass
	15 V	14.9483	14.9920	15.0519	0.0240	2.2	pass
	8 V	7.9724	7.9957	8.0277	0.0130	2.1	pass
	4 V	3.9862	3.9978	4.0138	0.0065	2.1	pass
	2 V	1.9931	1.9991	2.0069	0.0032	2.2	pass
	1 V	0.99655	0.99956	1.00346	0.00160	2.2	pass
	500 mV	498.28	499.78	501.73	0.81	2.1	pass
	250 mV	249.14	249.88	250.86	0.40	2.2	pass
	5 mV	4.9828	4.9979	5.0173	0.0100	1.7	pass
	2 V	1.9931	1.9995	2.0069	0.0030	2.3	pass
	500 kHz BW						
Channel 2	160 V	159.448	159.910	160.554	0.260	2.1	pass
	100 V	99.655	99.943	100.346	0.160	2.2	pass
	50 V	49.828	49.972	50.173	0.081	2.1	pass
	25 V	24.914	24.986	25.086	0.040	2.2	pass
	15 V	14.9483	14.9911	15.0519	0.0240	2.2	pass
	8 V	7.9724	7.9953	8.0277	0.0130	2.1	pass
	4 V	3.9862	3.9976	4.0138	0.0065	2.1	pass
	2 V	1.9931	1.9987	2.0069	0.0032	2.2	pass
	1 V	0.99655	0.99935	1.00346	0.00160	2.2	pass
	500 mV	498.28	499.67	501.73	0.81	2.1	pass
	250 mV	249.14	249.83	250.86	0.40	2.2	pass
	5 mV	4.9828	4.9968	5.0173	0.0100	1.7	pass
	2 V	1.9931	1.9998	2.0069	0.0030	2.3	pass
	500 kHz BW						
<b>[20] Level Meter AC Flatness, 1.15 Vrms (dB) - note 6</b>							
Channel 1	10 Hz	-0.0300	-0.0011	0.0300	0.0030	≥10	pass
	15 Hz	-0.0300	-0.0005	0.0300	0.0030	≥10	pass
	20 Hz	-0.0080	-0.0004	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0022	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0082	0.0300	0.0031	9.7	pass
	80 kHz	-0.0800	-0.0108	0.0800	0.0061	≥10	pass
	100 kHz	-0.0800	-0.0030	0.0800	0.0070	≥10	pass
	300 kHz	-0.200	-0.009	0.200	0.028	7.1	pass
	500 kHz BW						
	Channel 2	10 Hz	-0.0300	-0.0011	0.0300	0.0030	≥10
15 Hz		-0.0300	-0.0006	0.0300	0.0030	≥10	pass
20 Hz		-0.0080	-0.0005	0.0080	0.0029	2.8	pass
20 kHz		-0.0080	0.0019	0.0080	0.0024	3.3	pass
50 kHz		-0.0300	0.0079	0.0300	0.0031	9.7	pass
80 kHz		-0.0800	-0.0111	0.0800	0.0061	≥10	pass
100 kHz		-0.0800	-0.0027	0.0800	0.0070	≥10	pass
300 kHz		-0.200	-0.008	0.200	0.028	7.1	pass
500 kHz BW							

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
<b>ANALOG ANALYZER, continued</b>							
<b>[21] THD+N Harmonic Measurement Accuracy (dBV) - note 6</b>							
Channel 1 20 Hz notch	40 Hz	-20.300	-20.208	-19.700	0.011	≥10	pass
	60 Hz	-20.300	-20.070	-19.700	0.011	≥10	pass
	20 kHz	-20.300	-20.006	-19.700	0.011	≥10	pass
	100 kHz	-20.300	-20.009	-19.700	0.022	≥10	pass
Channel 1 1 kHz notch	20 Hz	-20.300	-20.002	-19.700	0.011	≥10	pass
	500 Hz	-20.300	-20.210	-19.700	0.011	≥10	pass
	2.0 kHz	-20.300	-20.206	-19.700	0.011	≥10	pass
	3.0 kHz	-20.300	-20.070	-19.700	0.011	≥10	pass
	20 kHz	-20.300	-20.007	-19.700	0.011	≥10	pass
	100 kHz	-20.300	-20.009	-19.700	0.022	≥10	pass
Channel 1 20 kHz notch	20 Hz	-20.300	-20.001	-19.700	0.011	≥10	pass
	10 kHz	-20.300	-20.209	-19.700	0.011	≥10	pass
	40 kHz	-20.300	-20.201	-19.700	0.011	≥10	pass
	60 kHz	-20.300	-20.066	-19.700	0.013	≥10	pass
	100 kHz	-20.300	-20.024	-19.700	0.022	≥10	pass
Channel 1 100 kHz notch	20 Hz	-20.300	-20.002	-19.700	0.011	≥10	pass
	20 kHz	-20.300	-20.028	-19.700	0.011	≥10	pass
	50 kHz	-20.300	-20.216	-19.700	0.011	≥10	pass
	200 kHz	-20.500	-20.180	-19.500	0.045	≥10	pass
	300 kHz	-20.500	-20.052	-19.500	0.063	7.9	pass
Channel 2 20 Hz notch	40 Hz	-20.300	-20.209	-19.700	0.011	≥10	pass
	60 Hz	-20.300	-20.070	-19.700	0.011	≥10	pass
	20 kHz	-20.300	-20.006	-19.700	0.011	≥10	pass
	100 kHz	-20.300	-20.009	-19.700	0.022	≥10	pass
Channel 2 1 kHz notch	20 Hz	-20.300	-20.002	-19.700	0.011	≥10	pass
	500 Hz	-20.300	-20.211	-19.700	0.011	≥10	pass
	2.0 kHz	-20.300	-20.208	-19.700	0.011	≥10	pass
	3.0 kHz	-20.300	-20.071	-19.700	0.011	≥10	pass
	20 kHz	-20.300	-20.007	-19.700	0.011	≥10	pass
	100 kHz	-20.300	-20.009	-19.700	0.022	≥10	pass
Channel 2 20 kHz notch	20 Hz	-20.300	-20.002	-19.700	0.011	≥10	pass
	10 kHz	-20.300	-20.210	-19.700	0.011	≥10	pass
	40 kHz	-20.300	-20.202	-19.700	0.011	≥10	pass
	60 kHz	-20.300	-20.067	-19.700	0.013	≥10	pass
	100 kHz	-20.300	-20.025	-19.700	0.022	≥10	pass
Channel 2 100 kHz notch	20 Hz	-20.300	-20.002	-19.700	0.011	≥10	pass
	20 kHz	-20.300	-20.029	-19.700	0.011	≥10	pass
	50 kHz	-20.300	-20.217	-19.700	0.011	≥10	pass
	200 kHz	-20.500	-20.187	-19.500	0.045	≥10	pass
	300 kHz	-20.500	-20.056	-19.500	0.063	7.9	pass
<b>[22] Frequency Measurement Accuracy (uHz/Hz) - note 1</b>							
	10 kHz	-3.00	-0.09	3.00	0.54	5.6	pass
<b>[23] Input Residual Crosstalk at 20 kHz (dB) - non-accredited</b>							
Ch2 into Ch 1		-999	-146.0	-130.0	4.0	na	pass
Ch 1 into Ch 2		-999	-146.0	-130.0	4.0	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result	
<b>NON-ACCREDITED ANALOG CHARACTERISTICS</b>								
<b>[24] Phase Measurement Offset (Deg), DC Coupled - non-accredited, self-test</b>								
Ch1 - Ch 2	50Hz	-0.150	0.000	0.150	0.001	na	pass	
	200Hz	-0.150	0.000	0.150	0.001	na	pass	
	5kHz	-0.150	-0.001	0.150	0.004	na	pass	
	20kHz	-0.600	-0.004	0.600	0.016	na	pass	
	50kHz	-1.500	-0.011	1.500	0.040	na	pass	
<b>[25] Squarewave Risetime (usec)- non-accredited, oscilloscope referenced</b>								
	5 kHz, 1 Vpk	0.00	1.38	1.70	0.06	na	pass	
<b>[26] Squarewave Even Harmonic Content (dB) - non-accredited, self-test</b>								
	20 Hz, 1 Vpk	-999	-127.0	-100.0	4.0	na	pass	
	5 kHz, 1 Vpk	-999	-117.6	-100.0	2.0	na	pass	
	20 kHz, 1 Vpk	-999	-101.8	-90.0	1.0	na	pass	
<b>[27] DAC Sine THD+N (dB) - non-accredited, self-test</b>								
Channel 1 Balanced	20 Hz,26.6V,22k BW	-999	-112.6	-105.0	1.5	na	pass	
	1 kHz,26.6V,22k BW	-999	-112.0	-105.0	1.5	na	pass	
	5 kHz,26.6V,22k BW	-999	-110.6	-105.0	1.5	na	pass	
	20 kHz,26.6V,22k BW	-999	-113.1	-105.0	1.5	na	pass	
	Unbalanced	20 Hz, 2V, 22k BW	-999	-112.7	-105.0	2.0	na	pass
		1 kHz, 2V, 22k BW	-999	-111.9	-105.0	2.0	na	pass
		20 kHz, 2V, 22k BW	-999	-113.7	-105.0	1.5	na	pass
		1 kHz, 2V 80k BW	-999	-102.7	-95.0	1.0	na	pass
Channel 2 Balanced	20 Hz,26.6V,22k BW	-999	-112.7	-105.0	1.5	na	pass	
	1 kHz,26.6V,22k BW	-999	-112.0	-105.0	1.5	na	pass	
	5 kHz,26.6V,22k BW	-999	-110.8	-105.0	1.5	na	pass	
	20 kHz,26.6V,22k BW	-999	-113.2	-105.0	1.5	na	pass	
	Unbalanced	20 Hz, 2V, 22k BW	-999	-113.1	-105.0	2.0	na	pass
		1 kHz, 2V, 22k BW	-999	-111.6	-105.0	2.0	na	pass
		20 kHz, 2V, 22k BW	-999	-113.7	-105.0	1.5	na	pass
		1 kHz, 2V 80k BW	-999	-102.7	-95.0	1.0	na	pass
Channel 2 Unbalanced	20 kHz, 2V, 80k BW	-999	-101.0	-95.0	1.5	na	pass	
	20 kHz, 2V, 80k BW	-999	-101.0	-95.0	1.5	na	pass	
<b>[28] Analog Sine THD+N (dB) - non-accredited, self-test</b>								
Channel 1 Balanced	20 Hz,26.6V,22k BW	-999	-120.7	-115.0	0.8	na	pass	
	1 kHz,26.6V,22k BW	-999	-120.7	-115.0	0.8	na	pass	
	5 kHz,26.6V,22k BW	-999	-120.1	-115.0	0.8	na	pass	
	20 kHz,26.6V,22k BW	-999	-121.3	-115.0	1.0	na	pass	
	Unbalanced	20 Hz, 2V, 22k BW	-999	-121.7	-116.5	0.8	na	pass
		1 kHz, 2V, 22k BW	-999	-121.6	-116.5	0.8	na	pass
		20 kHz, 2V, 22k BW	-999	-122.0	-116.5	1.0	na	pass
		20 kHz, 2V, 80k BW	-999	-116.3	-110.5	1.0	na	pass
		20 kHz, 2V, 100k BW	-999	-115.1	-110.0	1.0	na	pass
		50 kHz, 2V, 250k BW	-999	-109.9	-104.6	1.0	na	pass
		100 kHz, 2V, 500k BW	-999	-107.0	-101.6	1.0	na	pass
		200 kHz, 2V, 1M BW	-999	-102.6	-96.0	1.0	na	pass
		Channel 2 Balanced	20 Hz,26.6V,22k BW	-999	-120.8	-115.0	0.8	na
1 kHz,26.6V,22k BW	-999		-120.6	-115.0	0.8	na	pass	
5 kHz,26.6V,22k BW	-999		-119.7	-115.0	0.8	na	pass	
20 kHz,26.6V,22k BW	-999		-121.0	-115.0	1.0	na	pass	
Unbalanced	20 Hz, 2V, 22k BW		-999	-121.6	-116.5	0.8	na	pass
	1 kHz, 2V, 22k BW		-999	-121.6	-116.5	0.8	na	pass
	20 kHz, 2V, 22k BW		-999	-122.3	-116.5	1.0	na	pass
	20 kHz, 2V, 80k BW		-999	-116.2	-110.5	1.0	na	pass
	20 kHz, 2V, 100k BW		-999	-115.0	-110.0	1.0	na	pass
	50 kHz, 2V, 250k BW		-999	-109.8	-104.6	1.0	na	pass
	100 kHz, 2V, 500k BW		-999	-106.9	-101.6	1.0	na	pass
	200 kHz, 2V, 1M BW		-999	-102.3	-96.0	1.0	na	pass
	<b>[29] Residual Crosstalk (dB), Output Related, 20kHz - non-accredited, self-test</b>							
Unbalanced, 13.3V, 20 kHz.	2 into 1	-999	-150.3	-130.0	2.0	na	pass	
	1 into 2	-999	-148.5	-130.0	2.0	na	pass	
Balanced, 26.6V, 20 kHz.	2 into 1	-999	-162.5	-130.0	4.0	na	pass	
	1 into 2	-999	-160.4	-130.0	4.0	na	pass	

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
<b>[30] Residual Noise (uVolts) - non-accredited, self-test</b>							
Balanced Input, inputs shorted.	Ch 1, CCIR unwt'd	0	0.88	1.00	0.06	na	pass
	Ch 2, CCIR unwt'd	0	0.87	1.00	0.06	na	pass
	Ch 1, 80 kHz BW	0	1.69	2.00	0.08	na	pass
	Ch 2, 80 kHz BW	0	1.69	2.00	0.08	na	pass
	Ch 1, 250 kHz BW	0	3.81	8.00	0.80	na	pass
	Ch 2, 250 kHz BW	0	3.56	8.00	0.80	na	pass
	Ch 1, 500 kHz BW	0	5.36	10.30	0.80	na	pass
Ch 2, 500 kHz BW	0	5.18	10.30	0.80	na	pass	
<b>[31] Residual SMPTE IMD (%), 4:1, 60Hz:7kHz - non-accredited, self-test</b>							
Balanced, 26.66 Vrms.	Ch 1	0%	0.00023%	0.00100%	0.00030%	na	pass
	Ch 2	0%	0.00026%	0.00100%	0.00030%	na	pass
<b>[32] Residual MOD IMD (%), 4:1, 60Hz:7kHz, d2 + d3 - non-accredited, self-test</b>							
Balanced, 26.66 Vrms.	Ch 1	0%	0.00013%	0.00080%	0.00020%	na	pass
	Ch 2	0%	0.00016%	0.00080%	0.00020%	na	pass
<b>[33] Residual DFD IMD (%), mean 19.5kHz, diff 1kHz - non-accredited, self-test</b>							
Bal 26.66Vrms d2+d3	Ch 1	0%	0.00004%	0.00032%	0.00012%	na	pass
	Ch 2	0%	0.00004%	0.00032%	0.00012%	na	pass
Bal 26.66Vrms d2 only	Ch 1	0%	0.00003%	0.00018%	0.00006%	na	pass
	Ch 2	0%	0.00002%	0.00018%	0.00006%	na	pass
<b>[34] Residual CCIF IMD (%), mean 12.5kHz, diff 1kHz - non-accredited, self-test</b>							
Bal 26.66Vrms d2 only	Ch 1	0%	0.00004%	0.00036%	0.00012%	na	pass
	Ch 2	0%	0.00004%	0.00036%	0.00012%	na	pass
<b>[35] Residual DIM IMD, DIM100 (%) - non-accredited, self-test</b>							
Balanced, 26.66Vrms	Ch 1	0%	0.00084%	0.00180%	0.00020%	na	pass
	Ch 2	0%	0.00085%	0.00180%	0.00020%	na	pass
<b>AES/EBU DIGITAL I/O CHARACTERISTICS</b>							
<b>[36] Output Voltage Accuracy (Volts) - non-accredited, oscilloscope referenced</b>							
balanced	8.00 Vpp	7.12	7.89	8.88	0.13	na	pass
	0.800 Vpp	0.640	0.791	0.960	0.014	na	pass
unbalanced	2.00 Vpp	1.820	1.968	2.180	0.031	na	pass
	0.200 Vpp	0.164	0.197	0.236	0.004	na	pass
<b>[37] Input Voltage Measurement Accuracy (Volts) - non-accredited, oscilloscope referenced</b>							
balanced	8.00 Vpp	7.58	8.07	8.43	0.13	na	pass
	0.800 Vpp	0.735	0.814	0.865	0.014	na	pass
unbalanced	2.00 Vpp	1.894	2.027	2.106	0.031	na	pass
	0.200 Vpp	0.184	0.203	0.216	0.004	na	pass
<b>[38] Variable Rise Time Accuracy (nsec) - non-accredited, oscilloscope referenced</b>							
unbalanced	25 nsec	20.5	25.6	29.5	2.0	na	pass
	100 nsec	88.0	99.4	112.0	3.0	na	pass
<b>[39] Normal Mode Noise Amplitude (Volts) - non-accredited, oscilloscope referenced</b>							
unbalanced	0.500 Vpp	0.425	0.494	0.575	0.028	na	pass
	0.100 Vpp	0.065	0.097	0.135	0.008	na	pass
<b>[40] Common Mode Sine Amplitude (Volts) - note 2</b>							
balanced	20.00 Vpp	17.95	19.96	22.05	0.16	≥10	pass
	1.200 Vpp	1.030	1.190	1.370	0.010	≥10	pass
<b>[41] Jitter Amplitude Accuracy at 500 Hz (nsec) - non-accredited, self-test</b>							
Generator	1159.82	1153.9	1159.7	1165.7	1.0	na	pass
	781.10	777.09	781.02	785.11	0.70	na	pass
	357.70	355.81	357.66	359.59	0.36	na	pass
Analyzer average reading	1159.82	1043	1173	1277	15	na	pass
	357.70	320.9	361.8	394.5	5.0	na	pass
<b>[42] Jitter Flatness (dB) - non-accredited, self-test</b>							
	100 Hz	-0.50	-0.03	0.50	0.10	na	pass
	200 Hz	-0.50	-0.02	0.50	0.06	na	pass
	1 kHz	-0.50	0.01	0.50	0.06	na	pass
	5 kHz	-0.50	0.01	0.50	0.06	na	pass
	20 kHz	-0.50	0.03	0.50	0.06	na	pass
	50 kHz	-0.50	0.08	0.50	0.06	na	pass
<b>[43] Residual Jitter, 700 Hz - 100 kHz BW (psec, pk) - non-accredited, self-test</b>							
	48k sample rate	0	252	600	120	na	pass
	96k sample rate	0	263	600	120	na	pass
	192k sample rate	0	305	600	120	na	pass

END OF REPORT