



Calibration Services
 9290 SW Nimbus Ave
 Beaverton, Oregon 97008
 USA



Calibration Report

Accredited Calibration

Report Number: XXXXXXXXXX-XXXXXX-X

Model: APx585
Serial Number: XXXXXXXXXX

Data Type: AS SHIPPED, NEW
Date of Cal: 3-Apr-2026

Program:
 APxCalibration.exe 24.01

Internal Module Status and Data			
AP Name	Description	Serial No.	Revision
BRIO	Backplane module	XXXXXX	600
BSL3	Clock module	XXXXXX	201
BPLX	Analog Output module	XXXXXX	701
BCAS	Analog Input module	XXXXXX	506
BAES	DIO module	XXXXXX	304

Explanatory notes to the last three columns of the calibration report

"MU" - The column labeled "MU" lists the expanded measurement uncertainties derived from equipment specifications, repeatability data, and other significant sources. These are stated at a minimum confidence of 95% using a coverage factor k=2 (except as appropriate) following the recommendations in ISO/IEC 98-3 *Guide to the expression of uncertainty in measurement (GUM:1995)*, BIPM JCGM 100:2008, and NIST Technical Note 1297.

"TUR" - The column labeled "TUR" lists the test uncertainty ratio calculated by dividing the lesser of the lower and upper reading tolerances by the 95% expanded measurement uncertainty. An entry of "na" indicates [1] the specified limits are one-sided, or [2] the performance characteristic is not accredited.

"Result" - The column labeled "Result" lists color-coded assessments that the observed characteristic is within its specified limits of performance. There are three possible indications:

pass -- The *READING* is within the specified upper and lower limits reduced by guard-bands equal to the 95% expanded measurement uncertainty. The probability or risk of false acceptance is very low, typically <0.2%.

uncertain -- The *READING* is within the specified upper and lower limits, but it is close to one of the limits by an amount that is less than the 95% expanded measurement uncertainty. The probability or risk of false acceptance is elevated.

>> FAIL << -- The *READING* is outside of the specified limit range.

Accredited measurements listed in the following pages correlate to Audio Precision's Scope of Accreditation as noted:

- note 1 - Frequency Measurement
- note 2 - AC Voltage Measurement
- note 3 - AC Flatness Measurement
- note 4 - Resistance Measurement
- note 5 - DC Voltage Measurement
- note 6 - AC Voltage Source and AC Flatness Source for testing AC Measurement Equipment
- note 7 - DC Voltage Source for testing DC Measurement Equipment

This report is valid only when accompanied by a signed Certificate of Calibration.

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG GENERATOR CHARACTERISTICS							
[1] Sine Frequency Accuracy (Hz) - note 1							
	10 kHz	9999.9700	10000.0039	10000.0300	0.0064	4.7	pass
[2] Sine Amplitude Accuracy, 1 kHz (Volts, mVolts) - note 2							
Channel 1	7.200 V	7.1752	7.2010	7.2249	0.0020	≥10	pass
	4.000 V	3.9862	4.0021	4.0138	0.0011	≥10	pass
	2.000 V	1.99310	2.00052	2.00692	0.00072	9.7	pass
	1.000 V	0.99655	0.99949	1.00346	0.00036	9.7	pass
	300 mV	298.97	300.16	301.04	0.13	8.1	pass
	12 mV	11.959	11.995	12.042	0.013	3.2	pass
Channel 2	7.200 V	7.1752	7.2017	7.2249	0.0020	≥10	pass
	4.000 V	3.9862	4.0033	4.0138	0.0011	≥10	pass
	2.000 V	1.99310	2.00134	2.00692	0.00072	9.7	pass
	1.000 V	0.99655	1.00005	1.00346	0.00036	9.7	pass
	300 mV	298.97	300.25	301.04	0.13	8.1	pass
	12 mV	11.959	12.002	12.042	0.013	3.2	pass
Channel 3	7.200 V	7.1752	7.2013	7.2249	0.0020	≥10	pass
	4.000 V	3.9862	4.0034	4.0138	0.0011	≥10	pass
	2.000 V	1.99310	2.00063	2.00692	0.00072	9.7	pass
	1.000 V	0.99655	0.99983	1.00346	0.00036	9.7	pass
	300 mV	298.97	300.25	301.04	0.13	8.1	pass
	12 mV	11.959	11.999	12.042	0.013	3.2	pass
Channel 4	7.200 V	7.1752	7.2009	7.2249	0.0020	≥10	pass
	4.000 V	3.9862	4.0028	4.0138	0.0011	≥10	pass
	2.000 V	1.99310	2.00057	2.00692	0.00072	9.7	pass
	1.000 V	0.99655	0.99975	1.00346	0.00036	9.7	pass
	300 mV	298.97	300.23	301.04	0.13	8.1	pass
	12 mV	11.959	11.998	12.042	0.013	3.2	pass
Channel 5	7.200 V	7.1752	7.2016	7.2249	0.0020	≥10	pass
	4.000 V	3.9862	4.0038	4.0138	0.0011	≥10	pass
	2.000 V	1.99310	2.00127	2.00692	0.00072	9.7	pass
	1.000 V	0.99655	0.99977	1.00346	0.00036	9.7	pass
	300 mV	298.97	300.28	301.04	0.13	8.1	pass
	12 mV	11.959	12.000	12.042	0.013	3.2	pass
Channel 6	7.200 V	7.1752	7.2015	7.2249	0.0020	≥10	pass
	4.000 V	3.9862	4.0029	4.0138	0.0011	≥10	pass
	2.000 V	1.99310	2.00038	2.00692	0.00072	9.7	pass
	1.000 V	0.99655	0.99961	1.00346	0.00036	9.7	pass
	300 mV	298.97	300.21	301.04	0.13	8.1	pass
	12 mV	11.959	11.996	12.042	0.013	3.2	pass
Channel 7	7.200 V	7.1752	7.2012	7.2249	0.0020	≥10	pass
	4.000 V	3.9862	4.0033	4.0138	0.0011	≥10	pass
	2.000 V	1.99310	2.00072	2.00692	0.00072	9.7	pass
	1.000 V	0.99655	0.99977	1.00346	0.00036	9.7	pass
	300 mV	298.97	300.25	301.04	0.13	8.1	pass
	12 mV	11.959	12.000	12.042	0.013	3.2	pass
Channel 8	7.200 V	7.1752	7.2012	7.2249	0.0020	≥10	pass
	4.000 V	3.9862	4.0029	4.0138	0.0011	≥10	pass
	2.000 V	1.99310	2.00060	2.00692	0.00072	9.7	pass
	1.000 V	0.99655	0.99994	1.00346	0.00036	9.7	pass
	300 mV	298.97	300.21	301.04	0.13	8.1	pass
	12 mV	11.959	12.000	12.042	0.013	3.2	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG GENERATOR, continued							
[3] Sine Flatness, 1.15 Vrms (dB) - note 3							
Channel 1	10 Hz	-0.0080	-0.0004	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	-0.0003	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	-0.0026	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0008	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	0.0117	0.1000	0.0092	≥10	pass
Channel 2	10 Hz	-0.0080	-0.0001	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	-0.0002	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	-0.0024	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0078	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	-0.0448	0.1000	0.0092	≥10	pass
Channel 3	10 Hz	-0.0080	-0.0001	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	-0.0002	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	-0.0025	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0003	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	0.0268	0.1000	0.0092	≥10	pass
Channel 4	10 Hz	-0.0080	-0.0001	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	-0.0003	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	-0.0026	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0001	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	0.0327	0.1000	0.0092	≥10	pass
Channel 5	10 Hz	-0.0080	-0.0003	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	-0.0003	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	-0.0022	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0058	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	-0.0340	0.1000	0.0092	≥10	pass
Channel 6	10 Hz	-0.0080	-0.0001	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	-0.0002	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	-0.0025	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0053	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	-0.0193	0.1000	0.0092	≥10	pass
Channel 7	10 Hz	-0.0080	-0.0004	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	-0.0003	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	-0.0025	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0011	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	0.0111	0.1000	0.0092	≥10	pass
Channel 8	10 Hz	-0.0080	-0.0001	0.0080	0.0018	4.4	pass
	20 Hz	-0.0080	-0.0002	0.0080	0.0017	4.7	pass
	20 kHz	-0.0080	-0.0022	0.0080	0.0019	4.2	pass
	50 kHz	-0.0300	-0.0052	0.0300	0.0040	7.5	pass
	80 kHz	-0.1000	-0.0022	0.1000	0.0092	≥10	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG GENERATOR, continued							
[4] DC Offset (mV), Unbal - note 5							
Channel 1	7.200V	-18.10	1.50	18.10	0.40	≥10	pass
	1.000V	-2.600	0.257	2.600	0.046	≥10	pass
	100mV	-0.350	0.027	0.350	0.024	≥10	pass
Channel 2	7.200V	-18.10	-2.50	18.10	0.40	≥10	pass
	1.000V	-2.600	-0.449	2.600	0.046	≥10	pass
	100mV	-0.350	-0.044	0.350	0.024	≥10	pass
Channel 3	7.200V	-18.10	-3.34	18.10	0.40	≥10	pass
	1.000V	-2.600	-0.595	2.600	0.046	≥10	pass
	100mV	-0.350	-0.058	0.350	0.024	≥10	pass
Channel 4	7.200V	-18.10	-0.39	18.10	0.40	≥10	pass
	1.000V	-2.600	-0.069	2.600	0.046	≥10	pass
	100mV	-0.350	-0.006	0.350	0.024	≥10	pass
Channel 5	7.200V	-18.10	-0.50	18.10	0.40	≥10	pass
	1.000V	-2.600	-0.087	2.600	0.046	≥10	pass
	100mV	-0.350	-0.008	0.350	0.024	≥10	pass
Channel 6	7.200V	-18.10	1.09	18.10	0.40	≥10	pass
	1.000V	-2.600	0.204	2.600	0.046	≥10	pass
	100mV	-0.350	0.021	0.350	0.024	≥10	pass
Channel 7	7.200V	-18.10	1.02	18.10	0.40	≥10	pass
	1.000V	-2.600	0.193	2.600	0.046	≥10	pass
	100mV	-0.350	0.021	0.350	0.024	≥10	pass
Channel 8	7.200V	-18.10	1.61	18.10	0.40	≥10	pass
	1.000V	-2.600	0.280	2.600	0.046	≥10	pass
	100mV	-0.350	0.030	0.350	0.024	≥10	pass
[5] Source Resistance Accuracy (Ω) - note 4							
50Ω Unbal	Ch 1	49.000	50.283	51.000	0.060	≥10	pass
	Ch 2	49.000	50.326	51.000	0.060	≥10	pass
	Ch 3	49.000	50.282	51.000	0.060	≥10	pass
	Ch 4	49.000	50.267	51.000	0.060	≥10	pass
	Ch 5	49.000	50.286	51.000	0.060	≥10	pass
	Ch 6	49.000	50.284	51.000	0.060	≥10	pass
	Ch 7	49.000	50.268	51.000	0.060	≥10	pass
	Ch 8	49.000	50.289	51.000	0.060	≥10	pass
100Ω Bal	Ch 1	99.00	100.68	101.00	0.10	≥10	pass
	Ch 2	99.00	100.68	101.00	0.10	≥10	pass
	Ch 3	99.00	100.60	101.00	0.10	≥10	pass
	Ch 4	99.00	100.63	101.00	0.10	≥10	pass
	Ch 5	99.00	100.51	101.00	0.10	≥10	pass
	Ch 6	99.00	100.44	101.00	0.10	≥10	pass
	Ch 7	99.00	100.43	101.00	0.10	≥10	pass
	Ch 8	99.00	100.47	101.00	0.10	≥10	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER CHARACTERISTICS							
[6] DC Measurement Accuracy (Volts, mVolts) - note 7							
Channel 1	+100 V	99.200	100.012	100.800	0.060	≥10	pass
	+30 V	29.758	30.003	30.242	0.018	≥10	pass
	+10 V	9.9200	10.0013	10.0800	0.0060	≥10	pass
	+3 V	2.9758	3.0004	3.0242	0.0018	≥10	pass
	+1 V	0.99200	1.00019	1.00800	0.00060	≥10	pass
	+300 mV	297.10	299.86	302.90	0.18	≥10	pass
	0 mV	-0.800	0.148	0.800	0.046	≥10	pass
	-300 mV	-302.90	-299.56	-297.10	0.18	≥10	pass
	-1 V	-1.00800	-0.99907	-0.99200	0.00060	≥10	pass
	-3 V	-3.0242	-2.9974	-2.9758	0.0018	≥10	pass
	-10 V	-10.0800	-9.9916	-9.9200	0.0060	≥10	pass
	-30 V	-30.242	-29.973	-29.758	0.018	≥10	pass
-100V	-100.800	-99.918	-99.200	0.060	≥10	pass	
Channel 2	+100 V	99.200	99.975	100.800	0.060	≥10	pass
	+30 V	29.758	29.991	30.242	0.018	≥10	pass
	+10 V	9.9200	9.9974	10.0800	0.0060	≥10	pass
	+3 V	2.9758	2.9993	3.0242	0.0018	≥10	pass
	+1 V	0.99200	0.99977	1.00800	0.00060	≥10	pass
	+300 mV	297.10	299.67	302.90	0.18	≥10	pass
	0 mV	-0.800	0.046	0.800	0.046	≥10	pass
	-300 mV	-302.90	-299.59	-297.10	0.18	≥10	pass
	-1 V	-1.00800	-0.99898	-0.99200	0.00060	≥10	pass
	-3 V	-3.0242	-2.9969	-2.9758	0.0018	≥10	pass
	-10 V	-10.0800	-9.9901	-9.9200	0.0060	≥10	pass
	-30 V	-30.242	-29.968	-29.758	0.018	≥10	pass
-100V	-100.800	-99.905	-99.200	0.060	≥10	pass	
Channel 3	+100 V	99.200	99.955	100.800	0.060	≥10	pass
	+30 V	29.758	29.985	30.242	0.018	≥10	pass
	+10 V	9.9200	9.9956	10.0800	0.0060	≥10	pass
	+3 V	2.9758	2.9986	3.0242	0.0018	≥10	pass
	+1 V	0.99200	0.99948	1.00800	0.00060	≥10	pass
	+300 mV	297.10	299.49	302.90	0.18	≥10	pass
	0 mV	-0.800	-0.121	0.800	0.046	≥10	pass
	-300 mV	-302.90	-299.73	-297.10	0.18	≥10	pass
	-1 V	-1.00800	-0.99910	-0.99200	0.00060	≥10	pass
	-3 V	-3.0242	-2.9971	-2.9758	0.0018	≥10	pass
	-10 V	-10.0800	-9.9902	-9.9200	0.0060	≥10	pass
	-30 V	-30.242	-29.970	-29.758	0.018	≥10	pass
-100V	-100.800	-99.903	-99.200	0.060	≥10	pass	
Channel 4	+100 V	99.200	99.958	100.800	0.060	≥10	pass
	+30 V	29.758	29.985	30.242	0.018	≥10	pass
	+10 V	9.9200	9.9959	10.0800	0.0060	≥10	pass
	+3 V	2.9758	2.9985	3.0242	0.0018	≥10	pass
	+1 V	0.99200	0.99946	1.00800	0.00060	≥10	pass
	+300 mV	297.10	299.52	302.90	0.18	≥10	pass
	0 mV	-0.800	-0.113	0.800	0.046	≥10	pass
	-300 mV	-302.90	-299.75	-297.10	0.18	≥10	pass
	-1 V	-1.00800	-0.99926	-0.99200	0.00060	≥10	pass
	-3 V	-3.0242	-2.9976	-2.9758	0.0018	≥10	pass
	-10 V	-10.0800	-9.9927	-9.9200	0.0060	≥10	pass
	-30 V	-30.242	-29.978	-29.758	0.018	≥10	pass
-100V	-100.800	-99.930	-99.200	0.060	≥10	pass	

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[6] DC Measurement Accuracy (Volts, mVolts), continued							
<i>Channel 5</i>	+100 V	99.200	99.947	100.800	0.060	≥10	pass
	+30 V	29.758	29.983	30.242	0.018	≥10	pass
	+10 V	9.9200	9.9952	10.0800	0.0060	≥10	pass
	+3 V	2.9758	2.9985	3.0242	0.0018	≥10	pass
	+1 V	0.99200	0.99958	1.00800	0.00060	≥10	pass
	+300 mV	297.10	299.67	302.90	0.18	≥10	pass
	0 mV	-0.800	0.059	0.800	0.046	≥10	pass
	-300 mV	-302.90	-299.55	-297.10	0.18	≥10	pass
	-1 V	-1.00800	-0.99897	-0.99200	0.00060	≥10	pass
	-3 V	-3.0242	-2.9971	-2.9758	0.0018	≥10	pass
	-10 V	-10.0800	-9.9907	-9.9200	0.0060	≥10	pass
	-30 V	-30.242	-29.970	-29.758	0.018	≥10	pass
<i>Channel 6</i>	+100 V	99.200	99.944	100.800	0.060	≥10	pass
	+30 V	29.758	29.980	30.242	0.018	≥10	pass
	+10 V	9.9200	9.9946	10.0800	0.0060	≥10	pass
	+3 V	2.9758	2.9983	3.0242	0.0018	≥10	pass
	+1 V	0.99200	0.99942	1.00800	0.00060	≥10	pass
	+300 mV	297.10	299.51	302.90	0.18	≥10	pass
	0 mV	-0.800	-0.153	0.800	0.046	≥10	pass
	-300 mV	-302.90	-299.82	-297.10	0.18	≥10	pass
	-1 V	-1.00800	-0.99956	-0.99200	0.00060	≥10	pass
	-3 V	-3.0242	-2.9986	-2.9758	0.0018	≥10	pass
	-10 V	-10.0800	-9.9950	-9.9200	0.0060	≥10	pass
	-30 V	-30.242	-29.985	-29.758	0.018	≥10	pass
<i>Channel 7</i>	+100 V	99.200	99.940	100.800	0.060	≥10	pass
	+30 V	29.758	29.980	30.242	0.018	≥10	pass
	+10 V	9.9200	9.9944	10.0800	0.0060	≥10	pass
	+3 V	2.9758	2.9982	3.0242	0.0018	≥10	pass
	+1 V	0.99200	0.99935	1.00800	0.00060	≥10	pass
	+300 mV	297.10	299.46	302.90	0.18	≥10	pass
	0 mV	-0.800	-0.150	0.800	0.046	≥10	pass
	-300 mV	-302.90	-299.76	-297.10	0.18	≥10	pass
	-1 V	-1.00800	-0.99923	-0.99200	0.00060	≥10	pass
	-3 V	-3.0242	-2.9975	-2.9758	0.0018	≥10	pass
	-10 V	-10.0800	-9.9918	-9.9200	0.0060	≥10	pass
	-30 V	-30.242	-29.974	-29.758	0.018	≥10	pass
<i>Channel 8</i>	+100 V	99.200	99.952	100.800	0.060	≥10	pass
	+30 V	29.758	29.984	30.242	0.018	≥10	pass
	+10 V	9.9200	9.9953	10.0800	0.0060	≥10	pass
	+3 V	2.9758	2.9986	3.0242	0.0018	≥10	pass
	+1 V	0.99200	0.99953	1.00800	0.00060	≥10	pass
	+300 mV	297.10	299.58	302.90	0.18	≥10	pass
	0 mV	-0.800	-0.051	0.800	0.046	≥10	pass
	-300 mV	-302.90	-299.68	-297.10	0.18	≥10	pass
	-1 V	-1.00800	-0.99918	-0.99200	0.00060	≥10	pass
	-3 V	-3.0242	-2.9975	-2.9758	0.0018	≥10	pass
	-10 V	-10.0800	-9.9917	-9.9200	0.0060	≥10	pass
	-30 V	-30.242	-29.973	-29.758	0.018	≥10	pass
-100V	-100.800	-99.917	-99.200	0.060	≥10	pass	

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[7] Input Common Mode Rejection (mV) - non-accredited							
Channel 1 (5V CM signal)	3.2V range, 200 Hz	0	0.008	1.581	0.071	na	pass
	3.2V range, 5 kHz	0	0.147	1.581	0.083	na	pass
	3.2V range, 20 kHz	0	0.57	1.58	0.18	na	pass
	10V range, 20 kHz	0	0.66	15.81	0.23	na	pass
	100V range, 20 kHz	0	1.80	28.12	0.29	na	pass
Channel 2 (5V CM signal)	3.2V range, 200 Hz	0	0.017	1.581	0.071	na	pass
	3.2V range, 5 kHz	0	0.179	1.581	0.083	na	pass
	3.2V range, 20 kHz	0	0.70	1.58	0.18	na	pass
	10V range, 20 kHz	0	0.63	15.81	0.23	na	pass
	100V range, 20 kHz	0	2.55	28.12	0.29	na	pass
Channel 3 (5V CM signal)	3.2V range, 200 Hz	0	0.010	1.581	0.071	na	pass
	3.2V range, 5 kHz	0	0.143	1.581	0.083	na	pass
	3.2V range, 20 kHz	0	0.56	1.58	0.18	na	pass
	10V range, 20 kHz	0	0.48	15.81	0.23	na	pass
	100V range, 20 kHz	0	0.50	28.12	0.29	na	pass
Channel 4 (5V CM signal)	3.2V range, 200 Hz	0	0.007	1.581	0.071	na	pass
	3.2V range, 5 kHz	0	0.106	1.581	0.083	na	pass
	3.2V range, 20 kHz	0	0.41	1.58	0.18	na	pass
	10V range, 20 kHz	0	0.67	15.81	0.23	na	pass
	100V range, 20 kHz	0	4.28	28.12	0.29	na	pass
Channel 5 (5V CM signal)	3.2V range, 200 Hz	0	0.013	1.581	0.071	na	pass
	3.2V range, 5 kHz	0	0.154	1.581	0.083	na	pass
	3.2V range, 20 kHz	0	0.60	1.58	0.18	na	pass
	10V range, 20 kHz	0	0.95	15.81	0.23	na	pass
	100V range, 20 kHz	0	2.28	28.12	0.29	na	pass
Channel 6 (5V CM signal)	3.2V range, 200 Hz	0	0.017	1.581	0.071	na	pass
	3.2V range, 5 kHz	0	0.175	1.581	0.083	na	pass
	3.2V range, 20 kHz	0	0.69	1.58	0.18	na	pass
	10V range, 20 kHz	0	0.96	15.81	0.23	na	pass
	100V range, 20 kHz	0	1.05	28.12	0.29	na	pass
Channel 7 (5V CM signal)	3.2V range, 200 Hz	0	0.012	1.581	0.071	na	pass
	3.2V range, 5 kHz	0	0.099	1.581	0.083	na	pass
	3.2V range, 20 kHz	0	0.38	1.58	0.18	na	pass
	10V range, 20 kHz	0	0.31	15.81	0.23	na	pass
	100V range, 20 kHz	0	2.90	28.12	0.29	na	pass
Channel 8 (5V CM signal)	3.2V range, 200 Hz	0	0.011	1.581	0.071	na	pass
	3.2V range, 5 kHz	0	0.141	1.581	0.083	na	pass
	3.2V range, 20 kHz	0	0.55	1.58	0.18	na	pass
	10V range, 20 kHz	0	0.83	15.81	0.23	na	pass
	100V range, 20 kHz	0	1.78	28.12	0.29	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[8] Level Meter AC Accuracy, 1 kHz (Volts, mVolts) - note 6							
Channel 1	100 V	99.655	99.978	100.346	0.056	6.2	pass
	30 V	29.897	29.994	30.104	0.017	6.2	pass
	9 V	8.9690	8.9983	9.0311	0.0050	6.3	pass
	3 V	2.9897	2.9996	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89987	0.90311	0.00050	6.3	pass
	300 mV	298.97	299.95	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9995	5.0173	0.0031	5.6	pass
Channel 2	100 V	99.655	99.953	100.346	0.056	6.2	pass
	30 V	29.897	29.986	30.104	0.017	6.2	pass
	9 V	8.9690	8.9960	9.0311	0.0050	6.3	pass
	3 V	2.9897	2.9989	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89966	0.90311	0.00050	6.3	pass
	300 mV	298.97	299.88	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9982	5.0173	0.0031	5.6	pass
Channel 3	100 V	99.655	99.943	100.346	0.056	6.2	pass
	30 V	29.897	29.983	30.104	0.017	6.2	pass
	9 V	8.9690	8.9953	9.0311	0.0050	6.3	pass
	3 V	2.9897	2.9986	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89958	0.90311	0.00050	6.3	pass
	300 mV	298.97	299.86	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9979	5.0173	0.0031	5.6	pass
Channel 4	100 V	99.655	99.959	100.346	0.056	6.2	pass
	30 V	29.897	29.987	30.104	0.017	6.2	pass
	9 V	8.9690	8.9966	9.0311	0.0050	6.3	pass
	3 V	2.9897	2.9989	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89965	0.90311	0.00050	6.3	pass
	300 mV	298.97	299.88	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9983	5.0173	0.0031	5.6	pass
Channel 5	100 V	99.655	99.942	100.346	0.056	6.2	pass
	30 V	29.897	29.982	30.104	0.017	6.2	pass
	9 V	8.9690	8.9954	9.0311	0.0050	6.3	pass
	3 V	2.9897	2.9985	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89957	0.90311	0.00050	6.3	pass
	300 mV	298.97	299.85	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9979	5.0173	0.0031	5.6	pass
Channel 6	100 V	99.655	99.963	100.346	0.056	6.2	pass
	30 V	29.897	29.989	30.104	0.017	6.2	pass
	9 V	8.9690	8.9971	9.0311	0.0050	6.3	pass
	3 V	2.9897	2.9992	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89978	0.90311	0.00050	6.3	pass
	300 mV	298.97	299.92	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9988	5.0173	0.0031	5.6	pass
Channel 7	100 V	99.655	99.943	100.346	0.056	6.2	pass
	30 V	29.897	29.983	30.104	0.017	6.2	pass
	9 V	8.9690	8.9956	9.0311	0.0050	6.3	pass
	3 V	2.9897	2.9986	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89958	0.90311	0.00050	6.3	pass
	300 mV	298.97	299.85	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9978	5.0173	0.0031	5.6	pass
Channel 8	100 V	99.655	99.949	100.346	0.056	6.2	pass
	30 V	29.897	29.984	30.104	0.017	6.2	pass
	9 V	8.9690	8.9960	9.0311	0.0050	6.3	pass
	3 V	2.9897	2.9988	3.0104	0.0017	6.2	pass
	0.9 V	0.89690	0.89964	0.90311	0.00050	6.3	pass
	300 mV	298.97	299.88	301.04	0.17	6.2	pass
	5 mV	4.9828	4.9982	5.0173	0.0031	5.6	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ANALOG ANALYZER, continued							
[9] Level Meter AC Flatness, 1.15 Vrms (dB) - note 6							
Channel 1	10 Hz	-0.0080	-0.0014	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0004	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0013	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0032	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0225	0.1000	0.0061	≥10	pass
Channel 2	10 Hz	-0.0080	-0.0014	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0004	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0016	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0051	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0176	0.1000	0.0061	≥10	pass
Channel 3	10 Hz	-0.0080	-0.0014	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0004	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0008	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0004	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0289	0.1000	0.0061	≥10	pass
Channel 4	10 Hz	-0.0080	-0.0014	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0004	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0013	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0032	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0220	0.1000	0.0061	≥10	pass
Channel 5	10 Hz	-0.0080	-0.0013	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0003	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0014	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0039	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0204	0.1000	0.0061	≥10	pass
Channel 6	10 Hz	-0.0080	-0.0013	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0003	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0014	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0028	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0238	0.1000	0.0061	≥10	pass
Channel 7	10 Hz	-0.0080	-0.0012	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0002	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0014	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0031	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0229	0.1000	0.0061	≥10	pass
Channel 8	10 Hz	-0.0080	-0.0013	0.0080	0.0030	2.7	pass
	20 Hz	-0.0080	-0.0002	0.0080	0.0029	2.8	pass
	20 kHz	-0.0080	0.0016	0.0080	0.0024	3.3	pass
	50 kHz	-0.0300	0.0039	0.0300	0.0031	9.7	pass
	80 kHz	-0.1000	-0.0207	0.1000	0.0061	≥10	pass
[10] Frequency Measurement Accuracy (uHz/Hz) - note 1							
	10 kHz	-3.00	-0.39	3.00	0.64	4.7	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
NON-ACCREDITED CHARACTERISTICS							
[11] Phase Measurement Offset (Deg) - non-accredited, self-test							
Ch1 - Ch 2	200 Hz	-0.200	0.000	0.200	0.001	na	pass
	5 kHz	-0.200	0.004	0.200	0.004	na	pass
	20 kHz	-0.800	0.017	0.800	0.016	na	pass
	50 kHz	-2.000	0.042	2.000	0.040	na	pass
Ch1 - Ch 3	200 Hz	-0.200	0.000	0.200	0.001	na	pass
	5 kHz	-0.200	0.000	0.200	0.004	na	pass
	20 kHz	-0.800	0.002	0.800	0.016	na	pass
	50 kHz	-2.000	0.009	2.000	0.040	na	pass
Ch1 - Ch 4	200 Hz	-0.200	0.000	0.200	0.001	na	pass
	5 kHz	-0.200	0.006	0.200	0.004	na	pass
	20 kHz	-0.800	0.026	0.800	0.016	na	pass
	50 kHz	-2.000	0.067	2.000	0.040	na	pass
Ch1 - Ch 5	200 Hz	-0.200	0.000	0.200	0.001	na	pass
	5 kHz	-0.200	0.005	0.200	0.004	na	pass
	20 kHz	-0.800	0.021	0.800	0.016	na	pass
	50 kHz	-2.000	0.055	2.000	0.040	na	pass
Ch1 - Ch 6	200 Hz	-0.200	0.000	0.200	0.001	na	pass
	5 kHz	-0.200	-0.005	0.200	0.004	na	pass
	20 kHz	-0.800	-0.020	0.800	0.016	na	pass
	50 kHz	-2.000	-0.050	2.000	0.040	na	pass
Ch1 - Ch 7	200 Hz	-0.200	0.000	0.200	0.001	na	pass
	5 kHz	-0.200	-0.001	0.200	0.004	na	pass
	20 kHz	-0.800	-0.004	0.800	0.016	na	pass
	50 kHz	-2.000	-0.010	2.000	0.040	na	pass
Ch1 - Ch 8	200 Hz	-0.200	0.000	0.200	0.001	na	pass
	5 kHz	-0.200	0.002	0.200	0.004	na	pass
	20 kHz	-0.800	0.008	0.800	0.016	na	pass
	50 kHz	-2.000	0.021	2.000	0.040	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
NON-ACCREDITED, continued							
[12] Sine THD+N (dB), 20 kHz BW - non-accredited, self-test							
Channel 1	20 Hz, 14.4V, Bal	-999	-108.3	-103.0	1.0	na	pass
	1 kHz, 14.4V, Bal	-999	-107.9	-103.0	1.0	na	pass
	5 kHz, 14.4V, Bal	-999	-107.5	-103.0	1.0	na	pass
	20 kHz, 14.4V, Bal	-999	-109.8	-103.0	1.5	na	pass
	20 Hz, 7.2V, Unbal	-999	-112.4	-103.0	1.0	na	pass
	20 kHz,7.2V, Unbal	-999	-113.0	-103.0	1.5	na	pass
Channel 2	20 Hz, 14.4V, Bal	-999	-108.1	-103.0	1.0	na	pass
	1 kHz, 14.4V, Bal	-999	-107.9	-103.0	1.0	na	pass
	5 kHz, 14.4V, Bal	-999	-108.0	-103.0	1.0	na	pass
	20 kHz, 14.4V, Bal	-999	-109.8	-103.0	1.5	na	pass
	20 Hz, 7.2V, Unbal	-999	-111.6	-103.0	1.0	na	pass
	20 kHz,7.2V, Unbal	-999	-113.2	-103.0	1.5	na	pass
Channel 3	20 Hz, 14.4V, Bal	-999	-107.9	-103.0	1.0	na	pass
	1 kHz, 14.4V, Bal	-999	-107.5	-103.0	1.0	na	pass
	5 kHz, 14.4V, Bal	-999	-107.3	-103.0	1.0	na	pass
	20 kHz, 14.4V, Bal	-999	-109.3	-103.0	1.5	na	pass
	20 Hz, 7.2V, Unbal	-999	-112.0	-103.0	1.0	na	pass
	20 kHz,7.2V, Unbal	-999	-112.2	-103.0	1.5	na	pass
Channel 4	20 Hz, 14.4V, Bal	-999	-108.1	-103.0	1.0	na	pass
	1 kHz, 14.4V, Bal	-999	-108.0	-103.0	1.0	na	pass
	5 kHz, 14.4V, Bal	-999	-107.7	-103.0	1.0	na	pass
	20 kHz, 14.4V, Bal	-999	-109.4	-103.0	1.5	na	pass
	20 Hz, 7.2V, Unbal	-999	-111.5	-103.0	1.0	na	pass
	20 kHz,7.2V, Unbal	-999	-112.2	-103.0	1.5	na	pass
Channel 5	20 Hz, 14.4V, Bal	-999	-107.9	-103.0	1.0	na	pass
	1 kHz, 14.4V, Bal	-999	-108.1	-103.0	1.0	na	pass
	5 kHz, 14.4V, Bal	-999	-108.0	-103.0	1.0	na	pass
	20 kHz, 14.4V, Bal	-999	-109.6	-103.0	1.5	na	pass
	20 Hz, 7.2V, Unbal	-999	-112.2	-103.0	1.0	na	pass
	20 kHz,7.2V, Unbal	-999	-113.0	-103.0	1.5	na	pass
Channel 6	20 Hz, 14.4V, Bal	-999	-107.6	-103.0	1.0	na	pass
	1 kHz, 14.4V, Bal	-999	-108.1	-103.0	1.0	na	pass
	5 kHz, 14.4V, Bal	-999	-108.1	-103.0	1.0	na	pass
	20 kHz, 14.4V, Bal	-999	-109.8	-103.0	1.5	na	pass
	20 Hz, 7.2V, Unbal	-999	-111.2	-103.0	1.0	na	pass
	20 kHz,7.2V, Unbal	-999	-113.2	-103.0	1.5	na	pass
Channel 7	20 Hz, 14.4V, Bal	-999	-108.2	-103.0	1.0	na	pass
	1 kHz, 14.4V, Bal	-999	-108.0	-103.0	1.0	na	pass
	5 kHz, 14.4V, Bal	-999	-107.7	-103.0	1.0	na	pass
	20 kHz, 14.4V, Bal	-999	-109.7	-103.0	1.5	na	pass
	20 Hz, 7.2V, Unbal	-999	-112.8	-103.0	1.0	na	pass
	20 kHz,7.2V, Unbal	-999	-112.9	-103.0	1.5	na	pass
Channel 8	20 Hz, 14.4V, Bal	-999	-107.2	-103.0	1.0	na	pass
	1 kHz, 14.4V, Bal	-999	-107.2	-103.0	1.0	na	pass
	5 kHz, 14.4V, Bal	-999	-106.9	-103.0	1.0	na	pass
	20 kHz, 14.4V, Bal	-999	-109.5	-103.0	1.5	na	pass
	20 Hz, 7.2V, Unbal	-999	-110.0	-103.0	1.0	na	pass
	20 kHz,7.2V, Unbal	-999	-112.7	-103.0	1.5	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
NON-ACCREDITED, continued							
[13] Residual Crosstalk (dB) - non-accredited, self-test							
<i>Unbalanced, 7.2V, 20 kHz. Worst case, one channel driven.</i>	Ch 1	-999	-119.5	-115.0	2.0	na	pass
	Ch 2	-999	-119.0	-115.0	2.0	na	pass
	Ch 3	-999	-120.0	-115.0	2.0	na	pass
	Ch 4	-999	-118.9	-115.0	2.0	na	pass
	Ch 5	-999	-119.5	-115.0	2.0	na	pass
	Ch 6	-999	-118.7	-115.0	2.0	na	pass
	Ch 7	-999	-119.6	-115.0	2.0	na	pass
	Ch 8	-999	-130.5	-115.0	2.0	na	pass
<i>Balanced, 14.4V, 20 kHz. Worst case, one channel driven.</i>	Ch 1	-999	-108.3	-100.0	1.6	na	pass
	Ch 2	-999	-107.8	-100.0	1.6	na	pass
	Ch 3	-999	-107.9	-100.0	1.6	na	pass
	Ch 4	-999	-108.0	-100.0	1.6	na	pass
	Ch 5	-999	-107.3	-100.0	1.6	na	pass
	Ch 6	-999	-107.8	-100.0	1.6	na	pass
	Ch 7	-999	-108.6	-100.0	1.6	na	pass
	Ch 8	-999	-110.7	-100.0	1.6	na	pass
[14] Residual Noise (uVolts) - non-accredited, self-test							
<i>Balanced Input, inputs shorted, 20 kHz BW.</i>	Ch 1	0	1.04	1.30	0.06	na	pass
	Ch 2	0	1.05	1.30	0.06	na	pass
	Ch 3	0	1.04	1.30	0.06	na	pass
	Ch 4	0	1.05	1.30	0.06	na	pass
	Ch 5	0	1.04	1.30	0.06	na	pass
	Ch 6	0	1.05	1.30	0.06	na	pass
	Ch 7	0	1.06	1.30	0.06	na	pass
	Ch 8	0	1.05	1.30	0.06	na	pass
[15] Residual SMPTE IMD (%) - non-accredited, self-test							
<i>Unbalanced, 7.200 Vrms. 4:1 Ratio 60Hz:7kHz</i>	Ch 1	0%	0.00056%	0.00250%	0.00030%	na	pass
	Ch 2	0%	0.00038%	0.00250%	0.00030%	na	pass
	Ch 3	0%	0.00061%	0.00250%	0.00030%	na	pass
	Ch 4	0%	0.00040%	0.00250%	0.00030%	na	pass
	Ch 5	0%	0.00041%	0.00250%	0.00030%	na	pass
	Ch 6	0%	0.00042%	0.00250%	0.00030%	na	pass
	Ch 7	0%	0.00046%	0.00250%	0.00030%	na	pass
	Ch 8	0%	0.00089%	0.00250%	0.00030%	na	pass
[16] Residual MOD IMD (%) - non-accredited, self-test							
<i>Unbalanced, 7.200 Vrms. 4:1 Ratio 60Hz:7kHz</i>	Ch 1	0%	0.00066%	0.00250%	0.00020%	na	pass
	Ch 2	0%	0.00041%	0.00250%	0.00020%	na	pass
	Ch 3	0%	0.00066%	0.00250%	0.00020%	na	pass
	Ch 4	0%	0.00056%	0.00250%	0.00020%	na	pass
	Ch 5	0%	0.00052%	0.00250%	0.00020%	na	pass
	Ch 6	0%	0.00048%	0.00250%	0.00020%	na	pass
	Ch 7	0%	0.00055%	0.00250%	0.00020%	na	pass
	Ch 8	0%	0.00097%	0.00250%	0.00020%	na	pass
[17] Residual DFD IMD (%) - non-accredited, self-test							
<i>Unbalanced, 7.200 Vrms. Mean = 19.5kHz Diff = 1kHz</i>	Ch 1	0%	0.00020%	0.00100%	0.00012%	na	pass
	Ch 2	0%	0.00018%	0.00100%	0.00012%	na	pass
	Ch 3	0%	0.00020%	0.00100%	0.00012%	na	pass
	Ch 4	0%	0.00017%	0.00100%	0.00012%	na	pass
	Ch 5	0%	0.00015%	0.00100%	0.00012%	na	pass
	Ch 6	0%	0.00017%	0.00100%	0.00012%	na	pass
	Ch 7	0%	0.00014%	0.00100%	0.00012%	na	pass
	Ch 8	0%	0.00020%	0.00100%	0.00012%	na	pass
[18] Digital Output Amplitude Accuracy (Volts) - non-accredited							
<i>Unbal, consumer</i>	0.5 Vpp	0.450	0.495	0.550	0.009	na	pass
<i>Unbal, professional</i>	1.0 Vpp	0.900	0.977	1.100	0.014	na	pass
<i>Balanced</i>	5.0 Vpp	4.500	5.029	5.500	0.077	na	pass

Item	Setting(s)	Lower Limit	READING	Upper Limit	MU	TUR	Result
ADVANCED DIGITAL I/O CHARACTERISTICS				>> Option Not Installed			
[19] Output Voltage Accuracy (Volts) - non-accredited, oscilloscope referenced							
[20] Input Voltage Measurement Accuracy (Volts) - non-accredited, oscilloscope referenced							
[21] Variable Rise Time Accuracy (nsec) - non-accredited, oscilloscope referenced							
[22] Normal Mode Noise Amplitude (Volts) - non-accredited, oscilloscope referenced							
[23] Common Mode Sine Amplitude (Volts) - note 2							
[24] Jitter Amplitude Accuracy at 500 Hz (nsec) - non-accredited, self-test							
[25] Jitter Flatness (dB) - non-accredited, self-test							
[26] Residual Jitter, 700 Hz - 100 kHz BW (psec, pk) - non-accredited, self-test							

END OF REPORT

CONFIDENTIAL